




KV99 COVID-19 Killing Surgical Mask

VFE > 99.9a% / Killing COVID-19 99.81%

Efficiency



Properties

Test	
Differential Pressure (mmH ₂ O)	5
Fluid Resistance (mmHg)	160
Flame Spread (Second)	5
Microbio Cleanliness (cfu)	< 6
Shelf Life (Year)	5
ASTM F2101 Level 3 US Medical Face Masks Standard	
EN14683 TYPE IIR EU Surgical Masks Standard	
Core Filtration Material - Made in Taiwan / Hong Kong	
Formula - Food Additives Approved by World Health Organization (WHO)	
OOH technology pass chemical safety standards for baby textile products on EN ISO 21084:2019 EN ISO 18254:2016 EN ISO 14184:2011 JIS L 1041 DIN EN ISO 17070:2015 64 LFBG B 82.02-08 EN ISO 14389: 2014 US CPSC-CH-C1001-09.4	
Production process is determined to be a non-hazardous process according to EU Dangerous Preparation Directive 1999/45/EC	

Viral Filtration Efficiency 99.9a% / Killing COVID-19 99.81%



KV99 Flat Mask



KV99 3D Mask



KV99 Cup Mask




▶ The first Face Masks proven to **kill COVID-19 (99.81%)**

▶ Combined with Viral Filtration Efficiency **VFE (>99.9a%)**

▶ Over **500 times more effective** than traditional masks

Primary Infection




Breathing : > 1,000 COVID-19 viruses

Coughing : > 1,000,000 COVID-19 viruses

Infection : Breathing in > 100 active COVID-19 viruses


Type of Mask	No of Virus Penetrating Mask	
	Breathing	Coughing
VFE 95%	>50	>50,000
VFE 99%	>10	>10,000
Curie	>1	>10

Secondary Infection



- Viruses can survive on the surface of traditional masks for 7 days
- Viruses can grow 180 times more after 4 hours of usage
- We get infected by touching masks when they are full of viruses
- We bring masks with viruses back home, and cause secondary infection in our home environment
- A COVID-19 killing mask can fully protect against secondary infection

How We Arrest and Kill COVID-19




Spike protein over coronavirus is “strong negative charged”.

The highly positive-charged fibers inside Curie's filtration material attracts and arrests “strong negative charged” protein spikes on virus and bacteria

Strong attraction force can tear off protein chain over envelope of coronavirus, and achieve COVID-19 killing effect.

Shortcoming of Traditional Mask



Meltblown cloth relies on electrostatic to capture bacteria and viruses. Surface proteins of bacteria and viruses are negative-charged, and the electrostatic itself is negative. The polarity of the two is the same and therefore it repels – inefficient in capturing microbials.

Electrostatic itself is unstable, and it is easy to dissipate and discharge due to high temperature and humidity. Few hours after the mask used, the filtration efficiency begins to decay.



Curie

KV99 Flat Mask

(3M Nexcare Alternative)



Article No	: KV99-1
Product Name	: Curie Disposable Surgical Face Mask
Standard Applied	: EN14683 Type IIR / ASTM Level 3 / YY0469-2011
Specification	: Flat, Ear-Loop Mask, 17.5cmX9.5cm
Material	: 34% Positive Charged Biohazard Filter 22% Meltblown Non-Woven / 44% Non-Woven Fabric

Main Performance:

1) Particulate Filtration Efficiency (PFE)	: > 99%
2) Bacterial Filtration Efficiency (BFE)	: > 99.99%
3) Viral Filtration Efficiency (VFE)	: > 99.9a%
4) Bacterial Killing Rate	: > 99.99%
5) H3N2 Killing Rate	: > 99.99%
6) COVID-19 Killing Rate	: 99.81%
7) Differential Pressure	: < 5mmH2O
8) Splash Resistance Pressure	: < 160mmHg
9) Microbial Cleanliness	: < 30cfu/g
10) Flame Spread	: < 5 Seconds
11) Shelf Life	: 5 Years
Packing Details	: 1 Carton X 40 Boxes X 50 Pieces
Carton Size	: 570mm X 390mm X 420mm
Gross Weight	: 9.2kg

KV99 3D Mask

(3M 1870 Alternative)



Article No	: KV99-2
Product Name	: Curie Disposable Foldable Respirator
Standard Applied	: EN149 FFP2 / EN14683 Type IIR / ASTM Level 3
Specification	: Foldable, 10.7cm X 16cm
Material	: 34% Positive Charged Biohazard Filter 22% Meltblown Non-Woven / 44% Non-Woven Fabric

Main Performance:

1) Particulate Filtration Efficiency (PFE)	: > 99%
2) Bacterial Filtration Efficiency (BFE)	: > 99.99%
3) Viral Filtration Efficiency (VFE)	: > 99.9a%
4) Bacterial Killing Rate	: > 99.99%
5) H3N2 Killing Rate	: > 99.99%
6) COVID-19 Killing Rate	: 99.81%
7) Differential Pressure	: < 5mmH2O
8) Splash Resistance Pressure	: < 160mmHg
9) Microbial Cleanliness	: < 30cfu/g
10) Flame Spread	: < 5 Seconds
11) Shelf Life	: 5 Years
Packing Details	: 1 Carton X 20 Boxes X 30 Pieces
Carton Size	: 645mm X 310mm X 300mm
Gross Weight	: 4.9kg

KV99 Cup Mask

(3M 1860 Alternative)



Article No	: KV99-3
Product Name	: Curie Disposable Cup Shaped Respirator
Standard Applied	: NIOSH N95 / EN149 FFP2
Specification	: Cup Shaped, 10.7cm X 16cm
Material	: 34% Positive Charged Biohazard Filter 22% Meltblown Non-Woven / 44% Non-Woven Fabric

Main Performance:

1) Particulate Filtration Efficiency (PFE)	: > 99%
2) Bacterial Filtration Efficiency (BFE)	: > 99.99%
3) Viral Filtration Efficiency (VFE)	: > 99.9a%
4) Bacterial Killing Rate	: > 99.99%
5) H3N2 Killing Rate	: > 99.99%
6) COVID-19 Killing Rate	: 99.81%
7) Differential Pressure	: < 5mmH2O
8) Splash Resistance Pressure	: < 160mmHg
9) Microbial Cleanliness	: < 30cfu/g
10) Flame Spread	: < 5 Seconds
11) Shelf Life	: 5 Years
Packing Details	: 1 Carton X 20 Boxes X 20 Pieces
Carton Size	: 670mm X 435mm X 320mm
Gross Weight	: 6.5kg

KV99 4-Layer Surgical Grade Respirator / Arrest + Kill

Ultra-Protection, Anti-viral & Anti-bacterial, Light & Airy, Long-Lasting

Viral Filtration Efficiency 99.9a% / Killing COVID-19 99.81%

Strong Protection

Highest medical-grade protection (PFE > 99%, VFE / BFE > 99.9a%), meeting ASTM F2101 Level 3 and EN14683 standard, for a peace-of-mind.

Antiviral & Antibacterial

Kills COVID-19 (99.81%), virus (>99.99%) and bacteria (>99.99%) in seconds, greatly reducing risk for 2nd infection.

Light & Airy

Ultra-high breathability for more comfortable mask experience, 50% more breathable vs comparable masks.

Long-Lasting

Humidity and temperature proof – Curie patented strong polycationic nanostructure maintain ultrafiltration and disinfection efficiency, it can remain BFE > 99.9a% after conditioning in 120°C for 48 hours. Effective period more than for 5 years under optimal storage conditions.

KV99 4-Layer Surgical Grade Respirator / Arrest + Kill

Qualifications and Certifications



US Patent Number
62988900



香港特別行政區政府
知識產權署
Intellectual Property Department
The Government of the Hong Kong
Special Administrative Region

HK Patent Number
32020008506.8



US FDA Establishment Registration and Device Listing - Registered Establishment Number
3016752122

US FDA Establishment Registration and Device Listing - Owner/Operator Number
10063209

US FDA Personal Protective Equipment Emergency Use Authorization (PPE EUA)
Authorized Imported



NIOSH N95 Respirator Approval No:

TC-84A-7228

TC-84A-6973

EU CE Disposable Medical Face Mask Class I / Type IIR - Registration Number

DE/CA05/MP-238321-2556-00

DE/CA05/MP-238321-2557-00



EU CE EN149:2001+A1:2009 - FPC Certificate No

CE-PC-200402-188-FPC-A

CE-PC-200402-188-01-9A

CE-PC-200402-188-FPC-B

CE-PC-200601-452-01-9A

CEC40159



AS/NZS 1716:2012 - Respiratory Protective Devices Licence No:

SMK40477



China Chamber of Commerce for Import and Export of Medicines and Health
Products, White List of Medical Devices and Supplies Companies

- Unified Social Credit Identifier

91430181MA4PHUE510

KV99 4-Layer Surgical Grade Respirator / Arrest + Kill



Australian Government
Department of Health
Therapeutic Goods Administration

Australian Register of Therapeutic Goods Certificate - ARTG Identifier

Medical Device Included Class 1

338879

337859

338878

338713

337857



Brazil National Health Surveillance Agency ANVISA - Registration Number

Medical Device Class 1 – 81702110005



Health
Canada

Canada Medical Device Establishment Licence - Licence Number

Class 1 Manufacturer – 14507



Korea Textile Inspection & Testing Institute Registration Number

SUQ20-00040_M1

SUQ20-00041_M1

SUQ20-00042_M1

SUQ20-00043_M1



Taiwan Ministry of Health and Welfare Medical Device Manufacturing No.

008382



ISO9001:2015

Certificate No: USA20Q40924R0M

ISO 14001:2015

Certificate No: USA20E40925R0M

EN ISO 13485:2016

Certificate No: CN20/42091



Viral Filtration Efficiency (VFE) in ASTM F2101

Proven that Curie technology can effectively **filter virus (>99.9a%)**



Bacterial Filtration Efficiency with Increased Delivery Challenge (BFE) in

ASTM F2101 and EN14683

Proven that Curie technology can effectively filter increased challenge of bacteria (99.8%)

KV99 4-Layer Surgical Grade Respirator / Arrest + Kill



Viral Filtration Efficiency (VFE) in ASTM F2101

Proven that Curie technology can effectively **filter virus (>99.9a%)**

Bacterial Filtration Efficiency (BFE) in ASTM F2101 and EN14683

Proven that Curie technology can effectively filter bacteria (99.9%), complying ASTM F2101 Level 3 and EN14683 Type IIR

Air Exchange Pressure in ASTM F2101 / EN14683

Proven that air exchange pressure of KV99 masks can comply with standards of ASTM F2101 Level 3 and EN14683 Type IIR



Microbial Cleanliness (Bioburden) of Medical Masks in EN14683

Proven that total bioburden of KV99 can comply with standard of EN14683 Type IIR (<6 cfu)

Synthetic Blood Penetration Pressure in ASTM F2101 / ASTM F1862 / EN14683

Proven that KV99 mask can effectively intercept fluid (160mmHg) from penetration, complying ASTM F2101 Level 3 and EN14683 Type IIR

Sub-Micron Particulate Filtration Efficiency (0.1µm PSL) in ASTM F2101 / ASTM F2299 / EN14683

Proven that KV99 masks can effectively filter 0.1µm sub-micron particulate (>99.8%), complying ASTM F2101 Level 3 and EN14683 Type IIR



Standard for the Flammability of Clothing Textiles - US 16 CFR Part 1610

Proven that KV99 masks can comply with Class 1 - Normal Flammability



Intertek's Tick Mark

Proven that KV99 masks were independently tested by Intertek. Safety, quality and performance of KV99 masks are endorsed by Intertek.



Determination of Antiviral Activity of Textile Products BS ISO 18184

Proven that Curie technology can effectively **kill COVID-19 Coronavirus** SARS-CoV-2 (>**99.81%**)



紡織產業綜合研究所
Taiwan Textile Research Institute

Sub-Micron Particulate Filtration Efficiency (0.1µm PSL) in ASTM F2101 / ASTM F2299 / EN14683

Proven that KV99 masks can effectively filter 0.1µm sub-micron particulate (>99%), complying ASTM F2101 Level 3 and EN14683 Type IIR

KV99 4-Layer Surgical Grade Respirator / Arrest + Kill



紡織產業綜合研究所
Taiwan Textile Research Institute

Air Exchange Pressure in ASTM F2101 / EN14683

Proven that air exchange pressure of KV99 masks can comply with standards of ASTM F2101 Level 3 and EN14683 Type IIR

Synthetic Blood Penetration Pressure in ASTM F2101 / ASTM F1862 / EN14683

Proven that KV99 mask can effectively intercept fluid (160mmHg) from penetration, complying ASTM F2101 Level 3 and EN14683 Type IIR



香港公開大學
THE OPEN UNIVERSITY
OF HONG KONG

Bacterial Filtration Efficiency (BFE) in ASTM F2101

Proven that Curie technology can effectively filter bacteria (>99%)



香港公開大學
THE OPEN UNIVERSITY
OF HONG KONG

Standard Guide for Accelerated Ageing of Sterile Barrier Systems for Medical Devices in ASTM F1980-16

Bacterial Filtration Efficiency (BFE) in ASTM F2101

Proven that Curie technology can effectively filter bacteria (>99%), after conditioning KV99 masks in 120°C for 48 hours, to simulate storing in room temperature for 5 years.



香港公開大學
THE OPEN UNIVERSITY
OF HONG KONG

Determination of Antibacterial Activity of Textile Products BS EN ISO 20743

Proven that Curie technology can effectively kill bacteria (>99%) · In less than 60 seconds.



广东省微生物分析检测中心
Guang Dong Detection Center Of Microbiology

Determination of Antiviral Activity of Textile Products BS ISO 18184

Proven that Curie technology can effectively kill Influenza A Virus Subtype H3N2 (>99.99%)

Determination of alkylphenols (AP) of Textile Products EN ISO 21084:2019

Proven that Alkylphenols (AP) is not detected from Curie technology

Detection and Determination of Alkylphenol Ethoxylates (APEO) of Textile Products EN ISO 18254:2016

Proven that Alkylphenol Ethoxylates (APEO) is not detected from Curie technology



BUREAU
VERITAS

Determination of Formaldehyde - Free and Hydrolysed Formaldehyde of Textile Products EN ISO 14184:2011 / JIS L 1041

Proven that Formaldehyde is not detected from Curie technology, and it reach safety level for Type 1 – Baby < 36 Months

KV99 4-Layer Surgical Grade Respirator / Arrest + Kill



BUREAU
VERITAS

Determination of Tetrachlorophenol-, Trichlorophenol-, Dichlorophenol-, Monochlorophenol-Isomers and Pentachlorophenol Content of Textile Products
DIN EN ISO 17070:2015 / 64 LFBG B 82.02-08 (Modified)

Proven that Chlorophenols Content and Ortho-Phenylphenol (OPP) are not detected from Curie technology

Determination of the Phthalate Content of Textile Products EN ISO 14389: 2014 /
US CPSC-CH-C1001-09.4

Proven that Phthalate Content are not detected from Curie technology

KV99 4-Layer Surgical Grade Respirator / Arrest + Kill

Curie Filtration Technology

- Curie's Filtration Technology is applicable to different materials. This viral and bacterial filtration technology is proven by multiple lab tests by renowned certification organizations;
- Curie Filtration Technology arrests and kills virus and bacteria; obtained US and Hong Kong patents;
- Wide-range of uses for combating COVID-19 pandemic: masks and other daily-use products, providing protection against viruses and bacteria;
- Curie is the first to commercialize and mass-produce this high-efficiency, cost-effective virus killing technology
- This patented technology combines the team's expertise in chemistry and applied materials, cumulating in this "World 1st" technology.

Curie Advantage

The first application of Curie's Filtration Technology is applied to PPE mask products, and has been tested and proven to arrest and kill viruses and bacteria; "Ultra-Protection, Anti-viral & Anti-bacterial, Light & Airy, Long-Lasting", KV99 is a much better mask when compared to any other masks in the market.


	Disposable (and reusable) Masks Using Curie's Ultrahigh-Efficiency Filtration Tech	Traditional Disposable Mask Using Meltblown Material
Filtration / Killing Mechanism	<p>High efficiency: According to ASTM F2101 and EN14683 test results show: VFE > 99.9a% BFE > 99.99%</p> <p>The highly positive-charged fibers inside Curie's Filtration material attracts and arrests highly negative-charge protein spikes on virus and bacteria.</p>	<p>Meltblown cloth relies on electrostatic to capture bacteria and viruses. Surface proteins of bacteria and viruses are negative-charged, and the static electricity itself is negative. The polarity of the two is the same and therefore it repels – inefficient in capturing microbials.</p> <p>Electrostatic itself is unstable, and it is easy to dissipate and discharge due to high temperature and humidity. Few hours after the mask used, the filtration efficiency begins to decay.</p>
Principles of Protection	<p>Ultra-protection: According to the test performed by the Open University of Hong Kong, the extremely strong positive charge from the bottom layer of Curie's fibre can block virus and bacteria by tearing the outer envelope and membrane of virus and bacteria within 60 seconds. Killing Rate of COVID-19 – 99.81% Killing Rate of H3N2 > 99.99% Killing Rate of Bacteria > 99.99%</p>	<p>Traditional masks only act as a barrier to block bacteria or viruses from the user, but they do not have the function of killing. This means bacteria or viruses will start to multiply on the meltblown fabric fibres; therefore after few-hours of usage, the mask surface becomes infectious with bacteria and viruses, potentially causing secondary infection on whoever or whatever person or surface it comes into contact.</p>

KV99 4-Layer Surgical Grade Respirator / Arrest + Kill

Effective Period	<p>Long-lasting: Curie’s Filtration Technology material was tested against storage in an environment of 120°C for 48 hours, in accordance with ASTM F1980 standard simulating 5 year storage conditions.</p> <p>After the test, Curie’s filter materials still maintain high efficiency (BFE > 99.99%) and function. , It allows long term storage that large enterprises, governments or public institutions require.</p>	<p>The effective period is short, and it is most likely due to external factors in the transportation process (such as high temperature or improper storage), electrostatic loss of the meltblown is significant over-time, causing the product to be ineffective.</p>
Air Permeability	<p>High air permeability: Compared with traditional masks, the air resistance is reduced and the air permeability is high. Making a long period of usage more comfortable.</p>	<p>Most traditional masks sacrifice breathability in order to achieve the high filtration efficiency expected by users.</p>

KV99 4-Layer Surgical Grade Respirator / Arrest + Kill


Comparison between Curie's Mask and Traditional Disposable Mask Using Meltblown
Difference on Bacterial Filtration Efficiency (BFE) after Electrostatic Discharged




Curie's Mask
(BFE > 99%)

Traditional Disposable Mask Using Meltblown
(BFE < 90%)

Difference on Bacterial Growth after 8 Hours Use of Masks



Curie's Mask
(Killing Bacteria > 99%)



Traditional Disposable Mask Using Meltblown
(Full of Bacteria)

US Patent Number

US 62/988,900



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NUMBER	FILING or 371(c) DATE	GRP ART UNIT	FIL FEE REC'D	ATTY. DOCKET NO	TOT CLAIMS	IND CLAIMS
62/988,900	03/12/2020		85	WIPC2002		

166693
Law Offices of Sergei Orel, LLC
2125 Center Avenue, Suite 616
Fort Lee, NJ 07024

CONFIRMATION NO. 9401
UPDATED FILING RECEIPT



Date Mailed: 07/09/2020

Receipt is acknowledged of this provisional patent application. It will not be examined for patentability and will become abandoned not later than twelve months after its filing date. Any correspondence concerning the application must include the following identification information: the U.S. APPLICATION NUMBER, FILING DATE, NAME OF FIRST INVENTOR, and TITLE OF INVENTION. Fees transmitted by check or draft are subject to collection.

Please verify the accuracy of the data presented on this receipt. If an error is noted on this Filing Receipt, please submit a written request for a corrected Filing Receipt identifying the requested changes, preferably by including a properly marked-up ADS showing the changes with strike-through for deletions and underlining for additions. If you received a "Notice to File Missing Parts" or other Notice requiring a response for this application, please submit any request for correction to this Filing Receipt with your reply to the Notice. When the USPTO processes the reply to the Notice, the USPTO will generate another Filing Receipt incorporating the requested corrections provided that the request is grantable.

Inventor(s)

Jianliang GONG, Yau Ma Tei, HONG KONG;
Chun Yin OR, Cheung Sha Wan, HONG KONG;

Applicant(s)

Jianliang GONG, Yau Ma Tei, HONG KONG;
Chun Yin OR, Cheung Sha Wan, HONG KONG;

Power of Attorney:

Michael Yablonsky--40407

Permission to Access Application via Priority Document Exchange: Yes

Permission to Access Search Results: Yes

Applicant may provide or rescind an authorization for access using Form PTO/SB/39 or Form PTO/SB/69 as appropriate.

If Required, Foreign Filing License Granted: 04/02/2020

The country code and number of your priority application, to be used for filing abroad under the Paris Convention, is **US 62/988,900**

Projected Publication Date: None, application is not eligible for pre-grant publication

Non-Publication Request: No

Early Publication Request: No

**** MICRO ENTITY ****

Title

Air Filtration System and Manufacturing Method Therefor

Statement under 37 CFR 1.55 or 1.78 for AIA (First Inventor to File) Transition Applications: No

PROTECTING YOUR INVENTION OUTSIDE THE UNITED STATES

Since the rights granted by a U.S. patent extend only throughout the territory of the United States and have no effect in a foreign country, an inventor who wishes patent protection in another country must apply for a patent in a specific country or in regional patent offices. Applicants may wish to consider the filing of an international application under the Patent Cooperation Treaty (PCT). An international (PCT) application generally has the same effect as a regular national patent application in each PCT-member country. The PCT process **simplifies** the filing of patent applications on the same invention in member countries, **but does not result** in a grant of "an international patent" and does not eliminate the need of applicants to file additional documents and fees in countries where patent protection is desired.

Almost every country has its own patent law, and a person desiring a patent in a particular country must make an application for patent in that country in accordance with its particular laws. Since the laws of many countries differ in various respects from the patent law of the United States, applicants are advised to seek guidance from specific foreign countries to ensure that patent rights are not lost prematurely.

Applicants also are advised that in the case of inventions made in the United States, the Director of the USPTO must issue a license before applicants can apply for a patent in a foreign country. The filing of a U.S. patent application serves as a request for a foreign filing license. The application's filing receipt contains further information and guidance as to the status of applicant's license for foreign filing.

Applicants may wish to consult the USPTO booklet, "General Information Concerning Patents" (specifically, the section entitled "Treaties and Foreign Patents") for more information on timeframes and deadlines for filing foreign patent applications. The guide is available either by contacting the USPTO Contact Center at 800-786-9199, or it can be viewed on the USPTO website at <http://www.uspto.gov/web/offices/pac/doc/general/index.html>.

For information on preventing theft of your intellectual property (patents, trademarks and copyrights), you may wish to consult the U.S. Government website, <http://www.stopfakes.gov>. Part of a Department of Commerce initiative, this website includes self-help "toolkits" giving innovators guidance on how to protect intellectual property in specific countries such as China, Korea and Mexico. For questions regarding patent enforcement issues, applicants may call the U.S. Government hotline at 1-866-999-HALT (1-866-999-4258).

LICENSE FOR FOREIGN FILING UNDER

Title 35, United States Code, Section 184

Title 37, Code of Federal Regulations, 5.11 & 5.15

GRANTED

The applicant has been granted a license under 35 U.S.C. 184, if the phrase "IF REQUIRED, FOREIGN FILING LICENSE GRANTED" followed by a date appears on this form. Such licenses are issued in all applications where the conditions for issuance of a license have been met, regardless of whether or not a license may be required as

set forth in 37 CFR 5.15. The scope and limitations of this license are set forth in 37 CFR 5.15(a) unless an earlier license has been issued under 37 CFR 5.15(b). The license is subject to revocation upon written notification. The date indicated is the effective date of the license, unless an earlier license of similar scope has been granted under 37 CFR 5.13 or 5.14.

This license is to be retained by the licensee and may be used at any time on or after the effective date thereof unless it is revoked. This license is automatically transferred to any related applications(s) filed under 37 CFR 1.53(d). This license is not retroactive.

The grant of a license does not in any way lessen the responsibility of a licensee for the security of the subject matter as imposed by any Government contract or the provisions of existing laws relating to espionage and the national security or the export of technical data. Licensees should apprise themselves of current regulations especially with respect to certain countries, of other agencies, particularly the Office of Defense Trade Controls, Department of State (with respect to Arms, Munitions and Implements of War (22 CFR 121-128)); the Bureau of Industry and Security, Department of Commerce (15 CFR parts 730-774); the Office of Foreign Assets Control, Department of Treasury (31 CFR Parts 500+) and the Department of Energy.

NOT GRANTED

No license under 35 U.S.C. 184 has been granted at this time, if the phrase "IF REQUIRED, FOREIGN FILING LICENSE GRANTED" DOES NOT appear on this form. Applicant may still petition for a license under 37 CFR 5.12, if a license is desired before the expiration of 6 months from the filing date of the application. If 6 months has lapsed from the filing date of this application and the licensee has not received any indication of a secrecy order under 35 U.S.C. 181, the licensee may foreign file the application pursuant to 37 CFR 5.15(b).

SelectUSA

The United States represents the largest, most dynamic marketplace in the world and is an unparalleled location for business investment, innovation, and commercialization of new technologies. The U.S. offers tremendous resources and advantages for those who invest and manufacture goods here. Through SelectUSA, our nation works to promote and facilitate business investment. SelectUSA provides information assistance to the international investor community; serves as an ombudsman for existing and potential investors; advocates on behalf of U.S. cities, states, and regions competing for global investment; and counsels U.S. economic development organizations on investment attraction best practices. To learn more about why the United States is the best country in the world to develop technology, manufacture products, deliver services, and grow your business, visit <http://www.SelectUSA.gov> or call +1-202-482-6800.

HK Patent Number

32020008506.8

香港特別行政區政府知識產權署專利註冊處

Patents Registry, Intellectual Property Department

The Government of the Hong Kong Special Administrative Region



知識產權署

Intellectual Property Department

In reply please quote this ref.: 32020008506.8

Your ref.: C200101/UM1HK

Tel.: 3543 1464

Fax.: 2838 6315

eMode

08 June, 2020

WORLD IP CONSULTANCY

Unit B, 3/F, Cheong Yu Bldg

No. 143-151 Castle Peak Rd, Yuen Long

HONG KONG

Application for Grant of a Short-term Patent

Under Application No. 32020008506.8


We refer to your application for a short-term patent lodged on 02 June, 2020.


The above application is found to have satisfied the minimum requirements as laid down in section 114(2) of the Patents Ordinance. The accorded date of filing is 02 June, 2020.

In general, you will receive our further letter for the application at least five months after the date of this letter. If you do not receive any letter from us after this period, please contact us at 2961 6901.

This is a letter issued by Nigel LEE for Registrar of Patents.

(This is a computer-generated copy. No signature is required.)

 U.S. Department of Health & Human Services

 **U.S. FOOD & DRUG**
ADMINISTRATION

Follow FDA | En Español

SEARCH

HomeFoodDrugsMedical DevicesRadiation-Emitting ProductsVaccines, Blood & BiologicsAnimal & VeterinaryCosmeticsTobacco Products

Establishment Registration & Device Listing

FDA HomeMedical DevicesDatabases

1 result found for Owner Operator Name :
EEXI

New Search


Establishment Name	Registration Number	Current Registration Yr
HUNAN EEXI TECHNOLOGY&SERVICE CO.,LTD CHINA	3016752122	2020
<ul style="list-style-type: none">Respirator, Surgical - Inherent, Disposable Medical Face Mask, Model YX002, YX9001, YX9501; Inherent, Medical Face Mask, Model YX001, YX005, YX007		Manufacturer
<ul style="list-style-type: none">Accessory, Surgical Apparel - Inherent, Protective Suit, Model YXFFF001, YXFFF002		Manufacturer
<ul style="list-style-type: none">Face Mask (Except N95 Respirator) For General Public/Healthcare Personnel Per Iie Guidance - Disposable Medical Face Mask, Model: YX001, YX005, YX105, YX119; Disposable Medical Face Mask, Model: YX116; SURGICAL FACE MASK (NON-STERILE), Model: YX011; Surgical Face Mask (Non-Sterile), Model: YX118, YX011, YX121		Manufacturer

Can't find what you're looking for? [Try a new search](#)

Page Last Updated: 09/28/2020


Note: If you need help accessing information in different file formats, see Instructions for Downloading Viewers and Players.


Language Assistance Available: Español | 繁體中文 | Tiếng Việt | 한국어 | Tagalog | Русский | العربية | Kreyòl Ayisyen | Français | Polski | Português | Italiano | Deutsch | 日本語 | فارسی | English


 [Accessibility](#) | [Contact FDA](#) | [Careers](#) | [FDA Basics](#) | [FOIA](#) | [No FEAR Act](#) | [Nondiscrimination](#) | [Website Policies](#)

U.S. Food and Drug Administration
10903 New Hampshire Avenue
Silver Spring, MD 20993
Ph. 1-888-INFO-FDA (1-888-463-6332)

Combination Products
Advisory Committees
Science & Research

 U.S. Department of Health & Human Services

 U.S. Department of Health & Human Services

 **U.S. FOOD & DRUG**
ADMINISTRATION

Follow FDA | En Español

SEARCH

HomeFoodDrugsMedical DevicesRadiation-Emitting ProductsVaccines, Blood & BiologicsAnimal & VeterinaryCosmeticsTobacco Products

Establishment Registration & Device Listing

[FDA Home](#) [Medical Devices](#) [Databases](#)

[New Search](#) [Back To Search Results](#)

Establishment:
HUNAN EEXI TECHNOLOGY&SERVICE CO.,LTD
No.6, North Of Pingtuo Road, Liuyang Hi-Tech Industrial Development Zone
Liuyang Hunan, CN 410323
Registration Number: 3016752122
FEI Number*: 3016752122
Status: Active
Date Of Registration Status: 2020

Owner/Operator:
[Hunan EEXI Technology&Service Co.,Ltd](#)
No.6, North Of Pingtuo Road, Liuyang Hi-
Liuyang, Hunan CN 410323
Owner/Operator Number: [10063209](#)

Official Correspondent:
Zhiming Lin
No.6, North Of Pingtuo Road, Liuyang Hi-
Liuyang, Hunan CN 410323
Phone: 86-731-83371666

US Agent:
Yang LingYun
605 Sherwood Forest Place
Cary, NC US 27519
Phone: 919 4751044 Ext
Email: Usagency@126.Com

* Firm Establishment Identifier (FEI) should be used for identification of entities within the imports message set

Page Last Updated: 09/28/2020
Note: If you need help accessing information in different file formats, see Instructions for Downloading Viewers and Players.
Language Assistance Available: Español | 繁體中文 | Tiếng Việt | 한국어 | Tagalog | Русский | العربية | Kreyòl Ayisyen | Français | Polski | Português | Italiano | Deutsch | 日本語 | فارسی | English

<https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfRL/rl.cfm?rid=240648>



DEPARTMENT OF HEALTH & HUMAN SERVICES

NIOSH Reference: TN-19825
Mfr. Reference: GHCL-288

Centers for Disease Control
and Prevention (CDC)

National Institute for Occupational
Safety and Health (NIOSH)
National Personal Protective
Technology Laboratory (NPPTL)
P.O. Box 18070
Pittsburgh, PA 15236-0070
Phone: 412-386-4000
Fax: 412-386-4051
December 15, 2014

Mr. Wanghua Xu
Management Representative
Guangzhou Harley Commodity Company Limited
Floor 1 to 4, #8, Team 17, Sandong Village
Xinhua Street, Huadu District
Guangzhou 510800
CHINA

Dear Mr. Xu:

The National Institute for Occupational Safety and Health (NIOSH) has reviewed your request accepted September 17, 2014. This request was for approval of the model L-288 filtering facepiece air purifying respirator for protections against particulates at a N95 filter efficiency level, reference assembly matrix L-288 AMa.xls.

This request is granted. Approvals are granted for English language only on all documentation. It is the manufacturer's responsibility to correctly translate materials desired in languages other than English. Approval number TC-84A-7228 has been assigned. This respirator is approved for protection against particulates at a N95 filter efficiency level.


The CD enclosed with this letter contains the final respirator approval label. The abbreviated label has been accepted as submitted. The cautions and limitations which apply to this approval are on the approval label. Only those assemblies affected by this request, or where new approval numbers are assigned, apply to this approval action. Production approval labels cannot include information on unapproved configurations.

The approved assembly consists of the parts as listed on the approval label and the assembly matrix. Parts are to be marked with the numbers indicated on the approval label in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).

This certificate of approval is not an endorsement of the respirator by NIOSH, and such endorsement shall not be stated or implied in advertisements or other publicity. However, you may publicize the fact that this respirator has met the requirements of Title 42, *Code of Federal Regulations*, Part 84 (42 CFR 84).

No changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before changes are made.

Sincerely yours,

A handwritten signature in black ink, appearing to read 'Jonathan V. Szalajda', with a stylized flourish at the end.

Jonathan V. Szalajda
Acting Chief, Technology Evaluation Branch
National Personal Protective Technology Laboratory

Enclosures



TEST REPORT

Task Number: TN-19825

Manufacturer: Guangzhou Harley Commodity Company Limited

Prepared by: Nichole Petitta

Tests Conducted by: Nichole Petitta

Date: November 21, 2014

Respirator Tested: L-288 N95

Background Information

Respirator tested as per test request.

Tests Assigned

<u>Test Description</u>	<u>STP Number</u>
A. Exhalation Resistance Test	RCT-APR-STP-0003
B. Inhalation Resistance Test	TEB-APR-STP-0007
C. Sodium Chloride (NACL) N95 Test	TEB-APR-STP-0059

Overall Results

The respirator system tested met the requirements of all the procedures listed above.

**National Institute for Occupational Safety and Health
Respirator Branch
Test Data Sheet**



Task Number: TN-19825

Reference No.: CFR 84.180

Test: Exhalation Resistance Test

STP No.: 3

Manufacturer: Guangzhou Harley Commodity Company Limited

Filter Type: Filter Only

Item Tested: L-288 N95

Sample	Maximum Allowable Resistance (MM of H ₂ O)	Actual Resistance (MM of H ₂ O)	Result
	Exhalation	Exhalation	
1	25	10.8	PASS
2	25	10.7	PASS
3	25	10.3	PASS

Overall Result: PASS

Comments:

Samples were tested on Brooks meter 000283.

Was all equipment verified to be in calibration throughout all testing?



Yes



No

Signature:

Nichole L. Petitta

Date: 11/4/2014

Engineering Technician

National Institute for Occupational Safety and Health
Respirator Branch
Test Data Sheet



Task Number: TN-19825

Reference No.: CFR 84.180

Test: Inhalation Resistance Test

STP No.: 7

Manufacturer: Guangzhou Harley Commodity Company Limited

Item Tested: L-288 N95

Filter Type: Filter Only

Sample	Maximum Allowable Resistance (MM of H2O)	Actual Resistance (MM of H2O)	Result
	Inhalation	Inhalation	
1	35	11.4	PASS
2	35	11.7	PASS
3	35	10.9	PASS

Overall Result: PASS

Signature:

Nichole L. Petetta

Date: 11/4/2014

Engineering Technician

Task Number: TN-19825

Reference No.: CFR 84.180

Test: Inhalation Resistance Test

STP No.: 7

Manufacturer: Guangzhou Harley Commodity Company Limited

Item Tested: L-288 N95

Comments:

Samples were tested on Brooks meter 000036.

Was all equipment verified to be in calibration throughout all testing?



Yes



No

Signature:

Nichole L. Petitta

Date: 11/4/2014

Engineering Technician

National Institute for Occupational Safety and Health
Respirator Branch
Test Data Sheet



Task Number: TN-19825

Reference No.: CFR 84.181

Test: Sodium Chloride (NaCl) - N95

STP No.: 59

Manufacturer: Guangzhou Harley Commodity Company Limited

Item Tested: L-288 N95

Filter	Flow Rate	Initial Filter Resistance	Maximum Allowable Percent Leakage	Initial Percent Leakage	Maximum Percent Leakage	Result
1	85	12.2	5.00	0.379	0.388	PASS
2	85	12.2	5.00	0.559	0.560	PASS
3	85	11.6	5.00	0.781	0.765	PASS
4	85	12.0	5.00	0.650	0.650	PASS
5	85	11.7	5.00	0.665	0.665	PASS
6	85	12.3	5.00	0.520	0.520	PASS
7	85	10.9	5.00	0.598	0.598	PASS
8	85	12.5	5.00	1.390	1.490	PASS
9	85	11.3	5.00	0.794	0.794	PASS
10	85	13.2	5.00	0.654	0.654	PASS
11	85	13.3	5.00	0.610	0.610	PASS
12	85	11.4	5.00	0.629	0.629	PASS
13	85	12.1	5.00	0.640	0.640	PASS
14	85	12.0	5.00	0.755	0.755	PASS
15	85	11.7	5.00	0.899	0.934	PASS
16	85	12.2	5.00	0.616	0.616	PASS
17	85	12.0	5.00	0.684	0.684	PASS
18	85	12.7	5.00	0.501	0.501	PASS
19	85	11.0	5.00	0.319	0.326	PASS
20	85	11.4	5.00	0.669	0.669	PASS

Overall Result: PASS

Signature:

Nicholas L. Petetta

Date: 11/21/2014

Engineering Technician

Task Number: TN-19825

Reference No.: CFR 84.181

Test: Sodium Chloride (NaCl) - N95

STP No.: 59

Manufacturer: Guangzhou Harley Commodity Company Limited

Item Tested: L-288 N95

Comments:

Samples were tested on TSI 8130 machine 000192 with samples 1-3 using timer 000214.

Was all equipment verified to be in calibration throughout all testing?



Yes



No

Signature:

Nicholas L. Petitta

Date: 11/21/2014

Engineering Technician



DEPARTMENT OF HEALTH & HUMAN SERVICES

NIOSH Reference: TN-19824
Mfr. Reference: GHCL-188

Centers for Disease Control
and Prevention (CDC)

National Institute for Occupational
Safety and Health (NIOSH)
National Personal Protective
Technology Laboratory (NPPTL)
P.O. Box 18070
Pittsburgh, PA 15236-0070
Phone: 412-386-4000
Fax: 412-386-4051
December 15, 2014

Mr. Wanghua Xu
Management Representative
Guangzhou Harley Commodity Company Limited
Floor 1 to 4, #8, Team 17, Sandong Village
Xinhua Street, Huadu District
Guangzhou 510800
CHINA

Dear Mr. Xu:

The National Institute for Occupational Safety and Health (NIOSH) has reviewed your request accepted September 17, 2014. This request was for approval of the model L-188 N95 filtering facepiece air purifying respirator for protections against particulates at a N95 filter efficiency level, reference assembly matrix L-188 Ama.xls. In addition, the Guangzhou Harley Commodity Company QA Manual, version A1, effective date 2014-06-10, was submitted for review.

This request is granted. Approvals are granted for English language only on all documentation. It is the manufacturer's responsibility to correctly translate materials desired in languages other than English. Approval number TC-84A-6973 has been assigned. This respirator is approved for protection against particulates at a N95 filter efficiency level.

NIOSH has also reviewed the quality manual presented and finds that this manual meets or exceeds the minimum technical requirements for quality assurance plans as outlined in Title 42, *Code of Federal Regulations* (CFR), Part 84.41(a) and, on the basis of this review, this quality manual is accepted.

The CD enclosed with this letter contains the final respirator label. The abbreviated label is accepted as submitted. The cautions and limitations which apply to this approval are on the approval label. Only those assemblies affected by this request, or where new approval numbers are assigned, apply to this approval action. Production approval labels cannot include information on unapproved configurations.


The approved assembly consists of the parts as listed on the approval label and the assembly matrix. Parts are to be marked with the numbers indicated on the approval label in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).

This certificate of approval is not an endorsement of the respirator by NIOSH, and such endorsement shall not be stated or implied in advertisements or other publicity. However, you may publicize the fact that this respirator has met the requirements of Title 42, *Code of Federal Regulations*, Part 84 (42 CFR 84).

No changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before changes are made.


A copy of the quality manual will be retained by NIOSH and incorporated into our files. Any future changes to this accepted quality document must be submitted to NIOSH for a modification of this accepted quality system.

Sincerely yours,

A handwritten signature in black ink, appearing to read 'Jonathan V. Szalajda', with a stylized flourish at the end.

Jonathan V. Szalajda
Acting Chief, Technology Evaluation Branch
National Personal Protective Technology Laboratory

Enclosures



TEST REPORT

Task Number: TN-19824

Manufacturer: Guangzhou Harley Commodity Company Limited

Prepared by: Nichole Petitta

Tests Conducted by: Nichole Petitta

Date: November 13, 2014

Respirator Tested: L-188 N95

Background Information

Respirator tested as per test request.

Tests Assigned

<u>Test Description</u>	<u>STP Number</u>
A. Exhalation Resistance Test	RCT-APR-STP-0003
B. Inhalation Resistance Test	TEB-APR-STP-0007
C. Sodium Chloride (NACL) N95 Test	TEB-APR-STP-0059

Overall Results

The respirator system tested met the requirements of all the procedures listed above.

National Institute for Occupational Safety and Health
Respirator Branch
Test Data Sheet



Task Number: TN-19824

Reference No.: CFR 84.180

Test: Exhalation Resistance Test

STP No.: 3

Manufacturer: Guangzhou Harley Commodity Company Limited

Filter Type: Filter Only

Item Tested: L-188 N95

Sample	Maximum Allowable Resistance (MM of H ₂ O)	Actual Resistance (MM of H ₂ O)	Result
	Exhalation	Exhalation	
1	25	9.6	PASS
2	25	8.9	PASS
3	25	8.5	PASS

Overall Result: PASS

Comments:

Samples were tested on Brooks meter 0000283.

Was all equipment verified to be in calibration throughout all testing?

☒ Yes ☐ No

Signature:

Nichole L. Petitta

Date: 11/4/2014

Engineering Technician

National Institute for Occupational Safety and Health
Respirator Branch
Test Data Sheet



Task Number: TN-19824

Reference No.: CFR 84.180

Test: Inhalation Resistance Test

STP No.: 7

Manufacturer: Guangzhou Harley Commodity Company Limited

Item Tested: L-188 N95

Filter Type: Filter Only

Sample	Maximum Allowable Resistance (MM of H2O)	Actual Resistance (MM of H2O)	Result
	Inhalation	Inhalation	
1	35	10.5	PASS
2	35	10.5	PASS
3	35	9.7	PASS

Overall Result: PASS

Signature:

Nichole L. Petitta

Date: 11/4/2014

Engineering Technician

Task Number: TN-19824

Reference No.: CFR 84.180

Test: Inhalation Resistance Test

STP No.: 7

Manufacturer: Guangzhou Harley Commodity Company Limited

Item Tested: L-188 N95

Comments:

Samples were tested on Brooks meter 000036.

Was all equipment verified to be in calibration throughout all testing?

☒ Yes ☐ No

Signature:

Nichole L. Petitta

Date: 11/4/2014

Engineering Technician

National Institute for Occupational Safety and Health
Respirator Branch
Test Data Sheet



Task Number: TN-19824

Reference No.: CFR 84.181

Test: Sodium Chloride (NaCl) - N95

STP No.: 59

Manufacturer: Guangzhou Harley Commodity Company Limited

Item Tested: L-188 N95

Filter	Flow Rate	Initial Filter Resistance	Maximum Allowable Percent Leakage	Initial Percent Leakage	Maximum Percent Leakage	Result
1	85	9.6	5.00	0.517	0.517	PASS
2	85	10.6	5.00	0.694	0.694	PASS
3	85	9.8	5.00	0.718	0.718	PASS
4	85	10.3	5.00	0.651	0.654	PASS
5	85	9.3	5.00	0.682	0.682	PASS
6	85	9.7	5.00	0.755	0.755	PASS
7	85	10.7	5.00	0.621	0.621	PASS
8	85	9.9	5.00	0.744	0.744	PASS
9	85	9.8	5.00	0.535	0.535	PASS
10	85	10.8	5.00	1.500	1.580	PASS
11	85	10.5	5.00	0.688	0.688	PASS
12	85	9.7	5.00	0.692	0.692	PASS
13	85	10.6	5.00	0.757	0.757	PASS
14	85	10.8	5.00	0.671	0.671	PASS
15	85	10.8	5.00	0.981	0.981	PASS
16	85	10.3	5.00	0.678	0.678	PASS
17	85	9.8	5.00	0.661	0.661	PASS
18	85	10.3	5.00	0.680	0.680	PASS
19	85	9.6	5.00	1.660	1.660	PASS
20	85	9.8	5.00	0.769	0.769	PASS

Overall Result: PASS

Signature:

Nichole L. Petitta

Date: 11/13/2014

Engineering Technician

Task Number: TN-19824

Reference No.: CFR 84.181

Test: Sodium Chloride (NaCl) - N95

STP No.: 59

Manufacturer: Guangzhou Harley Commodity Company Limited

Item Tested: L-188 N95

Comments:

All samoles were tested on TSI 8130 machine 000192 with samples 1-3 using timer 000214.

Was all equipment verified to be in calibration throughout all testing?



Yes



No

Signature:

Nichole L. Petitta

Date: 11/13/2014

Engineering Technician

Appendix A: Authorized Imported, Non-NIOSH Approved Respirators Manufactured in China (Updated: September 28, 2020)

The table below includes a list of non-NIOSH respirators authorized by [this Umbrella EUA](#) for emergency use during the COVID-19 public health emergency.

As stated in the EUA, authorized respirators should be used in accordance with CDC's recommendations. For the most current CDC recommendations on optimizing respirator use, please visit CDC's webpage: [Strategies for Optimizing the Supply of N95 Respirators](#).

Search: Show entries

Manufacturer	Respirator Model(s)	Instructions for Use
Hunan EEXI Technology & Service Co., Ltd.	YX152, YX153	

Showing 1 to 1 of 1 entries (filtered from 166 total entries)

<https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/personal-protective-equipment-euas>

Allgemeine Anzeigepflicht nach §§ 25 und 30 Abs. 2 MPG **General Obligation to Notify pursuant to §§ 25 and 30 (2) Medical Devices Act, MPG**

Formblatt für Medizinprodukte, außer In-vitro-Diagnostika **Form for Medical Devices except In Vitro Diagnostic Medical Devices**

Zuständige Behörde / Competent authority	
Code DE/CA05	
Bezeichnung / Name Behörde für Gesundheit und Verbraucherschutz, Referat V43	
Staat / State Deutschland	Land / Federal state Hamburg
Ort / City Hamburg	Postleitzahl / Postal code 20539
Straße, Haus-Nr. / Street, house no. Billstraße 80	
Telefon / Phone +49-40-428280	Telefax / Fax +49-40-427310017
E-Mail / E-mail medizinprodukte@bgv.hamburg.de	

Anzeige / Notification	
Registriertdatum bei der zuständigen Behörde Registration date at competent authority 09.04.2020	Registriernummer / Registration number DE/CA05/MP-238321-2556-00
Typ der Anzeige / Notification type <input checked="" type="checkbox"/> Erstanzeige / Initial notification <input type="checkbox"/> Änderungsanzeige / Notification of change <input type="checkbox"/> Widerrufsanzeige / Notification of withdrawal	
Frühere Registriernummer bei Änderungs- und Widerrufsanzeige Previous registration number if notification has been changed or withdrawn	
Anzeigender nach § 25 MPG / Reporter pursuant to § 25 Medical Devices Act, MPG <input type="checkbox"/> Hersteller / Manufacturer <input checked="" type="checkbox"/> Bevollmächtigter / Authorised Representative <input type="checkbox"/> Einführer / Importer <input type="checkbox"/> Verantwortlicher für das Zusammensetzen von Systemen oder Behandlungseinheiten nach § 10 Abs. 1 und 2 MPG \ Assembler of systems or procedure packs pursuant to § 10 (1) and (2) Medical Devices Act, MPG <input type="checkbox"/> Betrieb oder Einrichtung (aufbereiten) nach § 25 Abs. 1 MPG i. V. m. § 4 Abs. 2 MPBetreibV Institution (processing) pursuant to § 25 (1) Medical Devices Act, MPG in connection with § 4 (2) MPBetreibV <input type="checkbox"/> Betrieb oder Einrichtung (sterilisieren) nach § 25 Abs. 2 i. V. m. § 10 Abs. 3 MPG Institution (sterilizing) pursuant to § 25 (2) in connection with § 10 (3) Medical Devices Act, MPG	

EC Declaration of Conformity

Manufacturer: Hunan EEXI Technology&Service Co.,Ltd.

Address: No.6, North of Pingtong road, Liuyang Hi-tech industrial development zone,
Hunan, China

EU Authorised Representative:

Shanghai International Holding Corp. GmbH (Europe)

Address: Eiffeistrasse 80, 20537 Hamburg, Germany

Device Name: Disposable medical face mask

Type: Type I

Classification : Class I

According to 93/42/EEC Annex IX, Rules 1, all non-invasive devices are in Class I, unless one of the rules set out hereinafter applies. The Conformity Route is Annex VII EC declaration of conformity.

We, Hunan EEXI Technology&Service Co.,Ltd. herewith declare on our exclusive responsibility that the above mentioned products meet the provisions of the Council Directive 93/42/EEC and 2007/47/EC for medical devices as transposed into national law. All supporting documentation is retained under the premises of the manufacturer.

The validity period of this declaration of conformity is limited by the issuing of a revised declaration of conformity after change of the product.

Harmonized Standards:

All applicable harmonized Standards (Published in the Official Journal of the European Communities)
Please see Annex List.

Date of Issue:

10th May 2020

Signature
Robbin Lin
General Manager
HUNAN EEXI TECHNOLOGY & SERVICE CO., LTD.
1610162282

Allgemeine Anzeigepflicht nach §§ 25 und 30 Abs. 2 MPG **General Obligation to Notify pursuant to §§ 25 and 30 (2) Medical Devices Act, MPG**

Formblatt für Medizinprodukte, außer In-vitro-Diagnostika **Form for Medical Devices except In Vitro Diagnostic Medical Devices**

Zuständige Behörde / Competent authority	
Code DE/CA05	
Bezeichnung / Name Behörde für Gesundheit und Verbraucherschutz, Referat V43	
Staat / State Deutschland	Land / Federal state Hamburg
Ort / City Hamburg	Postleitzahl / Postal code 20539
Straße, Haus-Nr. / Street, house no. Billstraße 80	
Telefon / Phone +49-40-428280	Telefax / Fax +49-40-427310017
E-Mail / E-mail medizinprodukte@bgv.hamburg.de	

Anzeige / Notification	
Registrierdatum bei der zuständigen Behörde Registration date at competent authority 09.04.2020	Registriernummer / Registration number DE/CA05/MP-238321-2557-00
Typ der Anzeige / Notification type <input checked="" type="checkbox"/> Erstanzeige / Initial notification <input type="checkbox"/> Änderungsanzeige / Notification of change <input type="checkbox"/> Widerrufsanzeige / Notification of withdrawal	
Frühere Registriernummer bei Änderungs- und Widerrufsanzeige Previous registration number if notification has been changed or withdrawn	
Anzeigender nach § 25 MPG / Reporter pursuant to § 25 Medical Devices Act, MPG <input type="checkbox"/> Hersteller / Manufacturer <input checked="" type="checkbox"/> Bevollmächtigter / Authorised Representative <input type="checkbox"/> Einführer / Importer <input type="checkbox"/> Verantwortlicher für das Zusammensetzen von Systemen oder Behandlungseinheiten nach § 10 Abs. 1 und 2 MPG \ Assembler of systems or procedure packs pursuant to § 10 (1) and (2) Medical Devices Act, MPG <input type="checkbox"/> Betrieb oder Einrichtung (aufbereiten) nach § 25 Abs. 1 MPG i. V. m. § 4 Abs. 2 MPBetreibV Institution (processing) pursuant to § 25 (1) Medical Devices Act, MPG in connection with § 4 (2) MPBetreibV <input type="checkbox"/> Betrieb oder Einrichtung (sterilisieren) nach § 25 Abs. 2 i. V. m. § 10 Abs. 3 MPG Institution (sterilizing) pursuant to § 25 (2) in connection with § 10 (3) Medical Devices Act, MPG	

EC Declaration of Conformity

Manufacturer: Hunan EEXI Technology&Service Co.,Ltd.

Address: No.6, North of Pingtong road, Liuyang Hi-tech industrial development zone,
Hunan, China

EU Authorised Representative:

Shanghai International Holding Corp. GmbH (Europe)

Address: Eiffestrasse 80, 20537 Hamburg, Germany

Device Name: Disposable Surgical mask

Type: Type IIR

Classification : Class I

According to 93/42/EEC Annex IX, Rules 1, all non-invasive devices are in Class I, unless one of the rules set out hereinafter applies. The Conformity Route is Annex VII EC declaration of conformity.

We, Hunan EEXI Technology&Service Co.,Ltd. herewith declare on our exclusive responsibility that the above mentioned products meet the provisions of the Council Directive 93/42/EEC and 2007/47/EC for medical devices as transposed into national law. All supporting documentation is retained under the premises of the manufacturer.

The validity period of this declaration of conformity is limited by the issuing of a revised declaration of conformity after change of the product.

Harmonized Standards:

All applicable harmonized Standards (Published in the Official Journal of the European Communities)
Please see Annex List.

Date of Issue:

10th May 2020

Signature:
Robin Lin
General Manager





Certificate of Module C2 production monitoring for equipment within the scope of Personal Protective Equipment Regulation (EU) 2016/425 Category III

FPC Certificate No.: CE-PC-200402-188-FPC-A

Certificate holder:	Hunan EEXI Technology & Service Co., Ltd. No.6, North of Pingtong Road, Liuyang Hi-Tech Industrial Development Zone, Hunan, China
The scope of the certification for:	Respiratory Protective Equipment Products covered by the certificate are described below.
Model:	Particle Filtering Half Mask YX009
Standard:	EN 149:2001+A1:2009
Validity from:	2020-05-29
To:	2020-08-28

CCQS Certification Services Limited in its role as a Notified Body for PPE Regulation, is monitoring that the manufacturer is producing PPE in conformity with the type described in the EU type-examination certificate and associated technical file and which satisfies the Essential Health and Safety Requirements of the Regulation. The manufacturer is hereby authorized to affix our Notified Body number, 2834, to each item of PPE as identified on this certificate whilst this certificate remains valid. This certificate is the property of CCQS and maybe withdrawn or revised at any time if CCQS considers that the equipment is no longer in conformity with the requirements of the Regulation.



Approved by Ireland
Government
as a Notified Body
for CE Marking No.2834



CCQS Certification Services Limited

Block 1 Blanchardstown Corporate Park, Ballycoolin Road, Blanchardstown, Dublin15,
D15 AKK1, Ireland

Tel: +00 353 1 588 6920 Website: www.ccqs.co.uk E-mail: info@ccqs.ie

If in any doubt about the integrity of this certificate, please contact CCQS by email to verify.



Module B EU Type-Examination Certificate

For the requirements of PPE Regulation 2016/425

Certificate No.: CE-PC-200402-188-01-9A

Certificate holder: Hunan EEXI Technology & Service Co., Ltd.
No.6, North of Pingtuo Road, Liuyang Hi-Tech Industrial Development Zone, Hunan, China

Product: **Particle Filtering Half Mask**
Folding filtering half mask without valve fitted with ear loops with head harness clip, external metal nose clip
Classification: FFP2 NR

Model reference: YX009

Standard(s): EN 149:2001+A1:2009

Test report No.: 2020(D) - 0179

Issue date: 2020-05-29

Expiry date: 2020-08-28

The product(s) on this certificate and the Technical File have been assessed and found to be in conformance with the Essential Health and Safety Requirements in Annex II of the PPE regulation 2016/425 and meeting the needs of WHO document dcp-ncov.pdf and EU Commission Recommendation (EU) 2020/403.

Any changes to the design, manufacturing location or manufacture of the PPE product certified here must be advised to CCQS Certification Services Limited for review.

CE marking shall not be applied until the requirements of all the PPE Regulation 2016/425 and relevant EN Harmonised standards and/or Technical specifications have been met.

If the certified product is Category III then this certificate is only valid if used in conjunction with Conformity Assessment against Module C2 or Module D.

This certificate remains the property of CCQS and maybe withdrawn at any time if it is considered that the equipment is no longer in conformity with the requirements of the PPE Regulation 2016/425.



Approved by Ireland
Government
as a Notified Body
for CE Marking No.2834



Approved by:

Owen Bian Director



CCQS Certification Services Limited

Block 1 Blanchardstown Corporate Park, Ballycoolin Road, Blanchardstown, Dublin15,
D15 AKK1, Ireland

Tel: +00 353 1 588 6920 Website: www.ccqs.co.uk E-mail: info@ccqs.ie

If in any doubt about the integrity of this certificate, please contact CCQS by email to verify.



Certificate of Module C2 production monitoring for equipment within the scope of Personal Protective Equipment Regulation (EU) 2016/425 Category III

FPC Certificate No.: CE-PC-200402-188-FPC-B

Certificate holder:	Hunan EEXI Technology & Service Co., Ltd. No.6, North of Pingtuo Road, Liuyang Hi-Tech Industrial Development Zone, Hunan, China
The scope of the certification for:	Respiratory Protective Equipment Products covered by the certificate are described below
Model:	Particle Filtering Half Mask YX009, YX135
Standard:	EN 149:2001+A1:2009
Validity from:	2020-05-29
Revision date	2020-06-19
To:	2020-08-28

CCQS Certification Services Limited in its role as a Notified Body for PPE Regulation, is monitoring that the manufacturer is producing PPE in conformity with the type described in the EU type-examination certificate and associated technical file and which satisfies the Essential Health and Safety Requirements of the Regulation. The manufacturer is hereby authorized to affix our Notified Body number, 2834, to each item of PPE as identified on this certificate whilst this certificate remains valid.
This certificate is the property of CCQS and maybe withdrawn or revised at any time if CCQS considers that the equipment is no longer in conformity with the requirements of the Regulation.



Approved by Ireland
Government
as a Notified Body
for CE Marking No.2834



Approved by:



CCQS Certification Services Limited

Block 1 Blanchardstown Corporate Park, Ballycoolin Road, Blanchardstown, Dublin15,
D15 AKK1, Ireland

Tel: +00 353 1 588 6920 Website: www.ccqs.co.uk E-mail: info@ccqs.ie

If in any doubt about the integrity of this certificate, please contact CCQS by email to verify.



Module B EU Type-Examination Certificate

For the requirements of PPE Regulation 2016/425

Certificate No.: CE-PC-200601-452-01-9A

Certificate holder: Hunan EEXI Technology & Service Co., Ltd.
No.6, North of Pingtuo Road, Liuyang Hi-Tech Industrial Development Zone, Hunan, China

Product: **Particle Filtering Half Mask**
Folding filtering half mask without valve fitted with ear loops with head harness clip, internal metal nose clip
Classification: FFP2 NR

Model reference: YX135

Standard(s): EN 149:2001+A1:2009

Test report No.: 2020(F) - 0033

Issue date: 2020-06-19

Expiry date: 2020-09-18

The product(s) on this certificate and the Technical File have been assessed and found to be in conformance with the Essential Health and Safety Requirements in Annex II of the PPE regulation 2016/425 and meeting the needs of WHO document dcp-ncov.pdf and EU Commission Recommendation (EU) 2020/403.

Any changes to the design, manufacturing location or manufacture of the PPE product certified here must be advised to CCQS Certification Services Limited for review.

CE marking shall not be applied until the requirements of all the PPE Regulation 2016/425 and relevant EN Harmonised standards and/or Technical specifications have been met.

If the certified product is Category III then this certificate is only valid if used in conjunction with Conformity Assessment against Module C2 or Module D.

This certificate remains the property of CCQS and maybe withdrawn at any time if it is considered that the equipment is no longer in conformity with the requirements of the PPE Regulation 2016/425.



Approved by Ireland
Government
as a Notified Body
for CE Marking No.2834



Approved by

Owen Bian - Director



CCQS Certification Services Limited

Block 1 Blanchardstown Corporate Park, Ballycoolin Road, Blanchardstown, Dublin15,
D15 AKK1, Ireland

Tel: +00 353 1 588 6920 Website: www.ccqs.co.uk E-mail: info@ccqs.ie

If in any doubt about the integrity of this certificate, please contact CCQS by email to verify.



EC TYPE EXAMINATION CERTIFICATE

SAI Global Assurance Services Limited ("SAI Global") (Notified Body No: **NB2056**) has examined the product and the related technical documentation as presented and certifies that the product complies with the Directive(s) and Standard or Specification as below and on the appendixes included with the certificate.

Certificate holder: **Guangzhou Harley Commodity Co., Ltd.**
Floor 1 to 4, #8, Team 17, Sandong Village, Xinhua Street,
Huadu District, Guangzhou, Guangdong,
P.R. China

Product Description: **Personal Protective Equipment (Category III)**
Respiratory Protective Devices

Examined for compliance with: **Directive PPE 89/686/EEC - (Article 10)**

Relevant Standard(s)/Technical Specification.

EN 149: 2001+A1:2009 - Filtering half masks to protect against particles

Original certification date: **17th Nov, 2015**

Date of issue: **17th Nov, 2015**

Expiry date: **16th Nov, 2020**

Certificate No: **CEC40159**

A handwritten signature in black ink, appearing to read 'Heather Mahon'.

Heather Mahon

Acting Head of Policy, Risk & Certification

Authorized Signatory

This certificate remains the property of SAI Global and has been issued in accordance with the CE Scheme Conditions and Procedures of the SAI Global Assurance Services Ltd. trading as EFSIS Ltd. Partis house, Ground Floor, Davy Avenue, Knowlhill. Milton Keynes, MK5 8HJ. United Kingdom. Notified Body: 2056

THIS CERTIFICATE DOES NOT ENTITLE THE HOLDER TO USE ANY OF THE CERTIFICATION TRADEMARKS ISSUED BY SAI GLOBAL LIMITED OR OTHERWISE.

APPENDIX 1

SAI Global Assurance Services Limited ("SAI Global") (Notified Body No: NB2056) has examined the product and the related technical documentation as presented and certifies that the product complies with the Directive(s) and Standard or Specification as below.

Product Description: **Respiratory Protective Devices (Category III)**

Series	Model	Description
L-210 series	L-210	Disposable particle half mask FFP1 NR, cup shaped, adjustable noseclip (polypropylene with iron type), velvet nose pad, two elastic headbands, non-valved
	L-210C	Disposable particle half mask FFP1 NR, with active carbon layer, cup shaped, adjustable noseclip (polypropylene with iron type), velvet nose pad, two elastic headbands, non-valved
	L-210V	Disposable particle half mask FFP1 NR, cup shaped, adjustable noseclip (polypropylene with iron type), velvet nose pad, two elastic headbands, valved with silicone flap
	L-210VC	Disposable particle half mask FFP1 NR, with active carbon layer, cup shaped, adjustable noseclip (polypropylene with iron type), velvet nose pad, two elastic headbands, valved with silicone flap
L-220 series	L-220	Disposable particle half mask FFP2 NR, cup shaped, adjustable noseclip (polypropylene with iron type), velvet nose pad, two elastic headbands, non-valved
	L-220C	Disposable particle half mask FFP2 NR, with active carbon layer, cup shaped, adjustable noseclip (polypropylene with iron type), velvet nose pad, two elastic headbands, non-valved
	L-220V	Disposable particle half mask FFP2 NR, cup shaped, adjustable noseclip (polypropylene with iron type), velvet nose pad, two elastic headbands, valved with silicone flap
	L-220VC	Disposable particle half mask FFP2 NR, with active carbon layer, cup shaped, adjustable noseclip (polypropylene with iron type), velvet nose pad, two elastic headbands, valved with silicone flap
L-201 series	L-201	Disposable particle half mask FFP1 NR, cup shaped, adjustable Aluminum noseclip, soft nosefoam, two elastic headbands, non-valved
	L-201C	Disposable particle half mask FFP1 NR, with active carbon layer, cup shaped, adjustable Aluminum noseclip, soft nosefoam, two elastic headbands, non-valved
	L-201V	Disposable particle half mask FFP1 NR, cup shaped, adjustable Aluminum noseclip, soft nosefoam, two elastic headbands, valved with silicone flap
	L-201VC	Disposable particle half mask FFP1 NR, with active carbon layer, cup shaped, adjustable Aluminum noseclip, soft nosefoam, two elastic headbands, valved with silicone flap

This certificate remains the property of SAI Global and has been issued in accordance with the CE Scheme Conditions and Procedures of the SAI Global Assurance Services Ltd. trading as EFSIS Ltd. Partis house, Ground Floor, Davy Avenue, Knowlhill. Milton Keynes, MK5 8HJ. United Kingdom. Notified Body: 2056

THIS CERTIFICATE DOES NOT ENTITLE THE HOLDER TO USE ANY OF THE CERTIFICATION TRADEMARKS ISSUED BY SAI GLOBAL LIMITED OR OTHERWISE.

APPENDIX 1 (CONT.)

Series	Model	Description
L-202 series	L-202	Disposable particle half mask FFP2 NR, cup shaped, adjustable Aluminum noseclip, soft nosefoam, two elastic headbands, non-valved
	L-202C	Disposable particle half mask FFP2 NR, with active carbon layer, cup shaped, adjustable Aluminum noseclip, soft nosefoam, two elastic headbands, non-valved
	L-202V	Disposable particle half mask FFP2 NR, cup shaped, adjustable Aluminum noseclip, soft nosefoam, two elastic headbands, valved with silicone flap
	L-202VC	Disposable particle half mask FFP2 NR, with active carbon layer, cup shaped, adjustable Aluminum nose clip, soft nosefoam, two elastic headbands, valved with silicone flap
L-203 series	L-203V	Disposable particle half mask FFP3 NR, cup shaped, adjustable nose clip, seal velvet, two elastic headbands with buckle, valved with silicone flap

Examined for compliance with: **Directive PPE 89/686/EEC (Article 10)**

Relevant Standard(s)/Technical Specification:

EN 149: 2001+A1:2009 - Filtering half masks to protect against particles

Technical file Reference No: **CEC40159 – CR26399**

Original date of issue: **17th Nov, 2015**

Date of issue: **17th Nov, 2015**

Certificate No: **CEC40159A1**



Heather Mahon

Acting Head of Policy, Risk & Certification

Authorized Signatory

This certificate remains the property of SAI Global and has been issued in accordance with the CE Scheme Conditions and Procedures of the SAI Global Assurance Services Ltd. trading as EFSIS Ltd. Partis house, Ground Floor, Davy Avenue, Knowlhill. Milton Keynes, MK5 8HJ. United Kingdom. Notified Body: 2056

THIS CERTIFICATE DOES NOT ENTITLE THE HOLDER TO USE ANY OF THE CERTIFICATION TRADEMARKS ISSUED BY SAI GLOBAL LIMITED OR OTHERWISE.

R-CES-13-1h Issue #10, July 2014



ARTICLE 11B Directive PPE 89/686/EEC CERTIFICATE

SAI Global Assurance Services Limited ("SAI Global") (Notified Body No: **NB2056**) confirm that the Management System of **Guangzhou Harley Commodity Co., Ltd.** complies with the requirements of Article 11b of the Directive PPE 89/686/EEC.

Certificate holder: **Guangzhou Harley Commodity Co., Ltd.**

Scope: **Manufacturing processes - Personal Protective Equipment:
Respiratory Protective Devices**

Relevant Standards:

EN 149: 2001+A1:2009 - Filtering half masks to protect against particles

Manufacturing sites/facilities covered by this certificate are:

**Guangzhou Harley Commodity Co., Ltd.
Floor 1 to 4, #8, Team 17, Sandong Village, Xinhua Street,
Huadu District, Guangzhou, Guangdong,
P.R. China**

Technical file Reference No: **CEC40159 – CR26399**

Original certification date: **17th Nov, 2015**

Date of issue: **17th Nov, 2015**

Expiry date: **16th Nov, 2020**

Certificate No: **CEC40159/11b**

Heather Mahon

Acting Head of Policy, Risk & Certification

Authorized Signatory

This certificate remains the property of SAI Global and has been issued in accordance with the CE Scheme Conditions and Procedures of the SAI Global Assurance Services Ltd. trading as EFSIS Ltd. Partis House, Ground Floor, Davy Avenue, Knowlhill. Milton Keynes, MK5 8HJ. United Kingdom.
Notified Body: NB2056

THIS CERTIFICATE DOES NOT ENTITLE THE HOLDER TO USE ANY OF THE CERTIFICATION TRADEMARKS ISSUED BY SAI GLOBAL LIMITED OR OTHERWISE.



STANDARDSMARK LICENCE



SCHEDULE TO STANDARDSMARK LICENCE

SAI Global hereby grants:

Guangzhou Harley Commodity Co., Ltd.

Floor 1 to 4, #8, Team 17, Sandong Village, Xinhua Street, Huadu District, Guangzhou, Guangdong China

StandardsMark Licence

Manufactured to:

AS/NZS 1716:2012 - Respiratory protective devices

Model identification of the goods on which the STANDARDSMARK may be used:

Model ID.	Face Piece / Head Covering Type	Filter or Cartridge Type/Class	Filter or Cartridge Model Number	Facepiece/Head covering Description	Model Name (Face Piece / Head Covering)	Size Facepiece/Head covering	Facepiece/Head covering-filtration type	Accessories or Comments	Date Endorsed
L-201	Half(1/2) facepiece-Disposable	P1	L-201	Cup shaped, adjustable Aluminum noseclip, soft nosefoam, two elastic headbands, non-valved			Filter/Cartridge(Work) Particle		6 Nov 2015
L-201C	Half(1/2) facepiece-Disposable	P1	L-201C	Cup shaped, with active carbon layer, adjustable Aluminum nose clip, soft nosefoam, two elastic headbands, non-valved			Filter/Cartridge(Work) Particle		6 Nov 2015
L-201V	Half(1/2) facepiece-Disposable	P1	L-201V	Cup shaped, adjustable Aluminum nose clip, soft nose foam, two elastic headbands, valved with silicone flap			Filter/Cartridge(Work) Particle		6 Nov 2015

Licence No: SMK40477

Issued Date: 9 November 2015

This schedule supersedes all previously issued schedules



* For details of manufacture, refer to the licensee

The STANDARDSMARK is a registered certification trademark of SAI Global Limited (A.C.N. 050 644 642) and is issued under licence by SAI Global Certification Services Pty Limited (ACN 108 716 669) ("SAI Global") 286 Sussex Street, Sydney NSW 2000, GPO Box 5420 Sydney NSW 2001. This certificate remains the property of SAI Global and must be returned to SAI Global upon its request. Refer to the Schedule for the list of product models.

SCHEDULE TO STANDARDSMARK LICENCE

Model ID.	Face Piece / Head Covering Type	Filter or Cartridge Type/Class	Filter or Cartridge Model Number	Facepiece/Head covering Description	Model Name (Face Piece / Head Covering)	Size Facepiece/Head covering	Facepiece/Head covering-filtration type	Accessories or Comments	Date Endorsed
L-201 VC	Half(1/2) facepiece-Disposable	P1	L-201VC	Cup shaped, with active carbon layer, adjustable Aluminum nose clip, soft nose foam, two elastic headbands, valved with silicone flap			Filter/Cartridge(Work) Particle		6 Nov 2015
L-202	Half(1/2) facepiece-Disposable	P2	L-202	Cup shaped, adjustable Aluminum nose clip, soft nose foam, two elastic headbands, non-valved			Filter/Cartridge(Work) Particle		6 Nov 2015
L-202 C	Half(1/2) facepiece-Disposable	P2	L-202C	Cup shaped, with active carbon layer, adjustable Aluminum nose clip, soft nose foam, two elastic headbands, non-valved			Filter/Cartridge(Work) Particle		6 Nov 2015
L-202 V	Half(1/2) facepiece-Disposable	P2	L-202V	Cup shaped, adjustable Aluminum nose clip, soft nose foam, two elastic headbands, valved with silicone flap			Filter/Cartridge(Work) Particle		6 Nov 2015
L-202 VC	Half(1/2) facepiece-Disposable	P2	L-202VC	Cup shaped, with active carbon layer, adjustable Aluminum nose clip, soft nose foam, two elastic headbands, valved with silicone flap			Filter/Cartridge(Work) Particle		6 Nov 2015

Licence No: SMK40477

Issued Date: 9 November 2015

This schedule supersedes all previously issued schedules



* For details of manufacture, refer to the licensee

The STANDARDSMARK is a registered certification trademark of SAI Global Limited (A.C.N. 050 644 642) and is issued under licence by SAI Global Certification Services Pty Limited (ACN 108 716 669) ("SAI Global") 286 Sussex Street, Sydney NSW 2000, GPO Box 5420 Sydney NSW 2001. This certificate remains the property of SAI Global and must be returned to SAI Global upon its request. Refer to the Schedule for the list of product models.

SCHEDULE TO STANDARDSMARK LICENCE

Model ID.	Face Piece / Head Covering Type	Filter or Cartridge Type/Class	Filter or Cartridge Model Number	Facepiece/Head covering Description	Model Name (Face Piece / Head Covering)	Size Facepiece/Head covering	Facepiece/Head covering-filtration type	Accessories or Comments	Date Endorsed
L-210	Half(1/2) facepiece-Disposable	P1	L-210	Cup shaped, adjustable nose clip (polypropylene with iron type), velvet nose pad, two elastic headbands, non-valved			Filter/Cartridge(Work) Particle		6 Nov 2015
L-210C	Half(1/2) facepiece-Disposable	P1	L-210C	Cup shaped, with active carbon layer, adjustable nose clip (polypropylene with iron type), velvet nose pad, two elastic headbands, non-valved			Filter/Cartridge(Work) Particle		6 Nov 2015
L-210V	Half(1/2) facepiece-Disposable	P1	L-210V	Cup shaped, adjustable nose clip (polypropylene with iron type), velvet nose pad, two elastic headbands, valved with silicone flap			Filter/Cartridge(Work) Particle		6 Nov 2015
L-210VC	Half(1/2) facepiece-Disposable	P1	L-210VC	Cup shaped, with active carbon layer, adjustable nose clip (polypropylene with iron type), velvet nose pad, two elastic headbands, valved with silicone flap			Filter/Cartridge(Work) Particle		6 Nov 2015

Licence No: SMK40477

Issued Date: 9 November 2015

This schedule supersedes all previously issued schedules



* For details of manufacture, refer to the licensee

The STANDARDSMARK is a registered certification trademark of SAI Global Limited (A.C.N. 050 644 642) and is issued under licence by SAI Global Certification Services Pty Limited (ACN 108 716 669) ("SAI Global") 286 Sussex Street, Sydney NSW 2000, GPO Box 5420 Sydney NSW 2001. This certificate remains the property of SAI Global and must be returned to SAI Global upon its request. Refer to the Schedule for the list of product models.

SCHEDULE TO STANDARDSMARK LICENCE

Model ID.	Face Piece / Head Covering Type	Filter or Cartridge Type/Class	Filter or Cartridge Model Number	Facepiece/Head covering Description	Model Name (Face Piece / Head Covering)	Size Facepiece/Head covering	Facepiece/Head covering-filtration type	Accessories or Comments	Date Endorsed
L-220	Half(1/2) facepiece-Disposable	P2	L-220	Cup shaped, adjustable nose clip (polypropylene with iron type), velvet nose pad, two elastic headbands, non-valved			Filter/Cartridge(Work) Particle		6 Nov 2015
L-220 C	Half(1/2) facepiece-Disposable	P2	L-220C	Cup shaped, with active carbon layer, adjustable nose clip (polypropylene with iron type), velvet nose pad, two elastic headbands, non-valved			Filter/Cartridge(Work) Particle		6 Nov 2015
L-220 V	Half(1/2) facepiece-Disposable	P2	L-220V	Cup shaped, adjustable nose clip (polypropylene with iron type), velvet nose pad, two elastic headbands, valved with silicone flap			Filter/Cartridge(Work) Particle		6 Nov 2015
L-220 VC	Half(1/2) facepiece-Disposable	P2	L-220V C	Cup shaped, with active carbon layer, adjustable nose clip (polypropylene with iron type), velvet nose pad, two elastic headbands, valved with silicone flap			Filter/Cartridge(Work) Particle		6 Nov 2015

End of Record

Licence No: SMK40477

Issued Date: 9 November 2015

This schedule supersedes all previously issued schedules



* For details of manufacture, refer to the licensee

The STANDARDSMARK is a registered certification trademark of SAI Global Limited (A.C.N. 050 644 642) and is issued under licence by SAI Global Certification Services Pty Limited (ACN 108 716 669) ("SAI Global") 286 Sussex Street, Sydney NSW 2000, GPO Box 5420 Sydney NSW 2001. This certificate remains the property of SAI Global and must be returned to SAI Global upon its request. Refer to the Schedule for the list of product models.

SCHEDULE TO STANDARDSMARK LICENCE

Licence No: SMK40477

Issued Date: 9 November 2015

This schedule supersedes all previously issued schedules



* For details of manufacture, refer to the licensee

The STANDARDSMARK is a registered certification trademark of SAI Global Limited (A.C.N. 050 644 642) and is issued under licence by SAI Global Certification Services Pty Limited (ACN 108 716 669) ("SAI Global") 286 Sussex Street, Sydney NSW 2000, GPO Box 5420 Sydney NSW 2001. This certificate remains the property of SAI Global and must be returned to SAI Global upon its request. Refer to the Schedule for the list of product models.

China Chamber of Commerce for Import and Export of Medicines and Health Products, White List of Medical Devices and Supplies Companies

- Unified Social Credit Identifier

91430181MA4PHUE510

① 不安全 | www.cccmhp.org.cn/Pub/6325/176241.shtml

首页 关于商会 新闻中心 行业服务 权威发布 商会会刊			
136	一顶湖北防护用品有限公司 One Top Protective Products Co., Ltd.	91429004316478369N	欧盟CE
137	常德葛澜医疗器械有限公司 Changde Gelan Medical Device Co.Ltd	91430700352838800T	欧盟CE
138	湖南佳龙医疗器械有限公司 Hunan Jialong Medical Equipment Co., Ltd	91431103MA4R4BF30T	欧盟CE
139	湖南特瑞精密医疗器械有限公司 Hunan Triplex Precision Medical Devices Co.,Ltd.	91430105MA4L3XD298	欧盟CE
140	湖南一喜科技服务有限公司 Hunan EEXI Technology&Service Co.,Ltd.	91430181MA4PHUE510	欧盟CE
141	湖南尤米服饰有限公司 Hunan Youmi Clothing Co., Ltd	91430181MA4LY6RK0Y	欧盟CE
142	常州美康纸塑制品有限公司 BH Medical Products Co., Ltd	91320400750008239J	澳大利亚 TGA
143	常州美康纸塑制品有限公司 BH Medical Products Co., Ltd	91320400750008239J	美国FDA
144	江西龙腾生物高科技有限公司 Jiangxi Longteng Biological High-Tech Co. Ltd	91360400741964050T	欧盟CE
145	江西美宝利医用敷料有限公司	91360426716524387R	欧盟CE
140	湖南一喜科技服务有限公司 Hunan EEXI Technology&Service Co.,Ltd.	91430181MA4PHUE510	欧盟CE

Australian Register of Therapeutic Goods Certificate - ARTG Identifier

Medical Device Included Class 1

338879



Australian Government

Department of Health
Therapeutic Goods Administration

Australian Register of Therapeutic Goods Certificate

Issued to

Auscanny Pty Ltd

for approval to supply

Auscanny Pty Ltd - Airway protection face mask

ARTG Identifier	338879
ARTG Start Date	3/07/2020
Product Category	Medical Device Included Class 1
GMDN	18094
GMDN Term	Airway protection face mask
Intended Purpose	Product Name: Particle filtering half mask Type: Folding filtering half mask without valve fitted with ear loops with head harness clip, internal metal nose clip. Intended Use: Single-use, disposable device, non-sterile, intended to be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations.

Manufacturer Details	Address	Certificate number(s)
Hunan EEXI Technology & Service Co Ltd	No 6 North of Pingtuo Road Liuyang Hi-tech Industrial Development Zone , Hunan , China	

ARTG Standard Conditions

The above Medical Device Included Class 1 has been entered on the Register subject to the following conditions:

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

Products Covered by This Entry

1. Airway protection face mask

Product Specific Conditions

No specific conditions have been recorded against this entry.

Therapeutic Goods Administration
PO Box 100, Woden ACT 2606 Australia
Phone: 1800 020 653
Email: info@tga.gov.au

ARTG Identifier: 338879
ARTG Start Date: 3/07/2020



Australian Government

Department of Health
Therapeutic Goods Administration

Australian Register of Therapeutic Goods Certificate

Issued to

Auscanny Pty Ltd

for approval to supply

Auscanny Pty Ltd - Airway protection face mask

ARTG Identifier	337859
ARTG Start Date	15/06/2020
Product Category	Medical Device Included Class 1
GMDN	18094
GMDN Term	Airway protection face mask
Intended Purpose	Single-use, disposable device, non-sterile, intended to be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations. This mask is made of blank mask, nose piece and tie-on belt. The blank mask is made of three layers. The inner and outer layer are non-woven fabric, the middle layer is Melt-blown polypropylene.

Manufacturer Details	Address	Certificate number(s)
Hunan EEXI Technology & Service Co Ltd	No 6 North of Pingtuo Road Liuyang Hi-tech Industrial Development Zone , Hunan , China	

ARTG Standard Conditions

The above Medical Device Included Class 1 has been entered on the Register subject to the following conditions:

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

Products Covered by This Entry

1. Airway protection face mask

Product Specific Conditions

No specific conditions have been recorded against this entry.

Therapeutic Goods Administration
PO Box 100, Woden ACT 2606 Australia
Phone: 1800 020 653
Email: info@tga.gov.au

ARTG Identifier: 337859
ARTG Start Date: 15/06/2020



Australian Government

Department of Health
Therapeutic Goods Administration

Australian Register of Therapeutic Goods Certificate

Issued to

Auscanny Pty Ltd

for approval to supply

Auscanny Pty Ltd - Airway protection face mask

ARTG Identifier	338878
ARTG Start Date	3/07/2020
Product Category	Medical Device Included Class 1
GMDN	18094
GMDN Term	Airway protection face mask
Intended Purpose	Product Name: Disposable medical face mask Type and Specification: Flat earloop, 14.5×9cm. Intended Use: Single-use, disposable device, non-sterile, intended to be used for children to reduce the risk of spread of infections particularly in epidemic or pandemic situations Structure and Components: This mask is made of blank mask, nose piece and ear loops. The blank mask is made of three layers. The inner and outer layers are non-woven fabric, the middle layer is melt-blown polypropylene.

Manufacturer Details	Address	Certificate number(s)
Hunan EEXI Technology & Service Co Ltd	No 6 North of Pingtuo Road Liuyang Hi-tech Industrial Development Zone , Hunan , China	

ARTG Standard Conditions

The above Medical Device Included Class 1 has been entered on the Register subject to the following conditions:

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

Products Covered by This Entry

1. Airway protection face mask

Product Specific Conditions

No specific conditions have been recorded against this entry.

Therapeutic Goods Administration
PO Box 100, Woden ACT 2606 Australia
Phone: 1800 020 653
Email: info@tga.gov.au

ARTG Identifier: 338878
ARTG Start Date: 3/07/2020



Australian Government

Department of Health
Therapeutic Goods Administration

Australian Register of Therapeutic Goods Certificate

Issued to

Auscanny Pty Ltd

for approval to supply

Auscanny Pty Ltd - Airway protection face mask

ARTG Identifier	338713
ARTG Start Date	1/07/2020
Product Category	Medical Device Included Class 1
GMDN	18094
GMDN Term	Airway protection face mask
Intended Purpose	Single-use, disposable device, non-sterile, intended to be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations. This mask is made of blank mask, nose piece and ear loops. The blank mask is made of three layers. The inner and outer layer are non-woven fabric, the middle layer is melt-blown polypropylene.

Manufacturer Details	Address	Certificate number(s)
Hunan EEXI Technology & Service Co Ltd	No 6 North of Pingtuo Road Liuyang Hi-tech Industrial Development Zone , Hunan , China	

ARTG Standard Conditions

The above Medical Device Included Class 1 has been entered on the Register subject to the following conditions:

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

Products Covered by This Entry

1. Airway protection face mask

Product Specific Conditions

No specific conditions have been recorded against this entry.

Therapeutic Goods Administration
PO Box 100, Woden ACT 2606 Australia
Phone: 1800 020 653
Email: info@tga.gov.au

ARTG Identifier: 338713
ARTG Start Date: 1/07/2020



Australian Government

Department of Health
Therapeutic Goods Administration

Australian Register of Therapeutic Goods Certificate

Issued to

Auscanny Pty Ltd

for approval to supply

Auscanny Pty Ltd - Airway protection face mask

ARTG Identifier	337857
ARTG Start Date	15/06/2020
Product Category	Medical Device Included Class 1
GMDN	18094
GMDN Term	Airway protection face mask
Intended Purpose	Single-use, disposable device, non-sterile, and intended to be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations. This mask is made of blank mask, nose piece and ear loops. The blank mask is made of three layers. The inner and outer layer are non-woven fabric, the middle layer is melt-blown polypropylene.

Manufacturer Details	Address	Certificate number(s)
Hunan EEXI Technology & Service Co Ltd	No 6 North of Pingtuo Road Liuyang Hi-tech Industrial Development Zone , Hunan , China	

ARTG Standard Conditions

The above Medical Device Included Class 1 has been entered on the Register subject to the following conditions:

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

Products Covered by This Entry

1. Airway protection face mask

Product Specific Conditions

No specific conditions have been recorded against this entry.

Therapeutic Goods Administration
PO Box 100, Woden ACT 2606 Australia
Phone: 1800 020 653
Email: info@tga.gov.au

ARTG Identifier: 337857
ARTG Start Date: 15/06/2020



MINISTÉRIO DA SAÚDE
AGÊNCIA NACIONAL DE VIGILÂNCIA SANITÁRIA

Este documento é para notificar que a fabricante

Hunan EEXI Technology&Service Co., Ltd.

CNPJ: 91430181MA4PHUE510

No.6, North of Pingtong road, Liuyang Hi-tech Industrial Development Zone, Hunan, China

Concluiu a notificação dos produtos a seguir perante à ANVISA – Agência Nacional de Vigilância Sanitária.

Materiais de uso médico das classes I.

NÚMERO DE REGISTRO: 81702110005

YX001 - Máscara cirúrgica descartável com elástico, tripla camada, 17,5x9,5cm

YX005 - Máscara cirúrgica descartável com elástico, tripla camada, 14,5x9cm

YX011 - Máscara cirúrgica descartável com elástico, tripla camada, 17,5x9,5cm

YX105 - Máscara cirúrgica descartável com elástico, tripla camada, 14,5x9cm

YX119 - Máscara cirúrgica descartável com tiras, tripla camada, 17,5x9,5cm

YX121 - Máscara cirúrgica descartável com tiras, tripla camada, 17,5x9,5cm

A autenticidade do registro pode ser conferida pelo seguinte procedimento:

1. Acesse <http://consultas.anvisa.gov.br/>
2. Clique em "Produtos para Saúde"
3. Em "Número do Registro", digite 81702110005
4. Clique em "consultar"
5. Clique no produto
6. Verifique os documentos notificados e a validade do registro (VIGENTE)

Licence Number

14507

Numéro de la licence

Medical Device
Establishment LicenceLicence d'établissement
pour les instruments médicaux**HUNAN EEXI TECHNOLOGY & SERVICE CO., LTD**#6 NORTH OF PINGTOU ROAD
LINYANG HI-TECH INDUSTRIAL DEVELOPMENT ZONE
LIUYANG, HUNAN
CHINA
410300This licence is issued in accordance with the Medical
Devices Regulations of the Food and Drugs Act for the
following activities:Cette licence est délivrée conformément à la Loi sur les
aliments et drogues, règlement sur les instruments
médicaux pour les activités qui suivent:

	Distributor / Distributeur	Importer / Importateur	Manufacture Devices for Distribution / Fabricant d'instruments médicaux pour distribution
Class I / Classe I	No / Non	No / Non	Yes / Oui
Class II / Classe II	No / Non	No / Non	
Class III / Classe III	No / Non	No / Non	
Class IV / Classe IV	No / Non	No / Non	

Attestation made :**Attestations faites :**

The establishment has documented procedures in place in respect of:		L'établissement a mis en oeuvre une procédure écrite concernant:
• distribution records	[Y]	• les registres de distribution
• complaint handling	[Y]	• les plaintes
• recalls	[Y]	• les rappels
• mandatory problem reporting	[N]	• rapports d'incident obligatoires
• handling, storage, delivery	[N]	• la manutention, le stockage, la livraison
• installation	[N]	• l'installation,
• corrective action	[N]	• les mesures correctives
• servicing	[N]	• l'entretien

Site listing begins on the back of this page**Liste des sites commence au verso de
cette page**

Issue Date, date de délivrance: 2020-07-17

Minister of Health Ministre de la santé	Countersigned: Director, Medical Devices Compliance Program or delegated authority Contresigné par: Directeur, Programme de la conformité des matériels médicaux ou autorité déléguée  Chad Sheehy
--	--

This licence is the property of the Medical Devices Compliance Program and must be returned upon demand.
Cette licence appartient au Programme de la conformité des matériels médicaux et doit être retournée sur demande.**Canada**



Health
Canada

Santé
Canada

Sites

HUNAN EEXI TECHNOLOGY & SERVICE CO., LTD
#6 NORTH OF PINGTOU ROAD
LINYANG HI-TECH INDUSTRIAL DEVELOPMENT ZONE
LIUYANG, HUNAN
CHINA
410300

Company ID

151728

No. d'identification
de la société

시험 성적서

KOTITI

접수번호 : 8220-1509-101001-001

발급번호 : 8135522286

접수일자 : 2020년 06월 09일

발급일자 : 2020년 06월 25일

시험명	KF94 Face Mask	검사목적	연구/참고/기타
제조번호	YX015	제조일자	-
증량/용량	-	유통(품질유지)기한	-
업체명	Hunan EEXI Technology & Service CO., Ltd	대표자	Lisa
소재지	No.6, North of Pingyou Road, Liuyang Hi-tech industrial development zone, Hunan, China	전화번호	86-15521119635

시험 항목 및 결과

시험항목	시험기준			단위	결과			비고
성상	겉면은 백색의 부직포이며, 가로 방향으로 3단으로 접혀있으며, 코 부위에는 코란이 있고, 양 측면에 탄력성이 있는 나이론 끈이 부착되어 있다.			-	겉면은 백색의 부직포이며, 가로 방향으로 3단으로 접혀있으며, 코 부위에는 코란이 있고, 양 측면에 탄력성이 있는 나이론 끈이 부착되어 있다.			자사기준 및 시험방법
형상	본체	가로	200 ± 10	mm	본체	가로	197	자사기준 및 시험방법
		세로	70 ± 10			세로	74	
	머리끈길이	좌	155 ± 15		머리끈길이	좌	141	
		우	155 ± 15			우	149	
순도시험 (색소)	위에서 관찰할 때 색을 나타내지 않는다.			-	위에서 관찰할 때 색을 나타내지 않는다.			의약외품에 관한 기준 및 시험방법
순도시험 (산 또는 알칼리)	홍색을 나타내지 않는다./ 적색을 나타내지 않는다.			-	홍색을 나타내지 않는다./ 적색을 나타내지 않는다.			의약외품에 관한 기준 및 시험방법
순도시험 (형광증백제)	형광을 나타내지 않아야 한다.			-	형광을 나타내지 않는다.			의약외품에 관한 기준 및 시험방법

시험검사자 : 조운섭, 권수지, 권아영

기술책임자 : 왕만식

※ 위 결과는 의뢰된 시험·검사 항목만을 대상으로 한 것입니다.

※ 지면이 부족한 경우 시험·검사 항목 및 결과란은 별지로 작성 가능합니다.

※ 검사결과를 광고하거나 용기·포장 등에 표시할 때에는 시험·검사성적서 전체 내용을 모두 표시하여야 합니다.

본 시험성적서는 「식품의약품분야 시험·검사 등에 관한 법률」에 따른 업무와 관련이 없으며, 이로 인한 법적 분쟁시 식품의약품안전처와 관련이 없음을 밝힙니다.

(사)KOTITI시험연구원



시험 성적서

KOTITI

접수번호 : 8219-1509-100779-001

발급번호 : 8680586806

접수일자 : 2020년 05월 15일

발급일자 : 2020년 06월 19일

시험 항목 및 결과

시험항목	시험기준	단위	결과	비고
순도시험 (포름알데히드)	검액이 나타내는 색은 비교액이 나타내는 색보다 진하지 않다.	-	검액이 나타내는 색은 비교액이 나타내는 색보다 진하지 않다.	의약품품에 관한 기준 및 시험방법
분진포집효율 (염화나트륨)	개개의 측정치는 94 %이상	%	본품 : 100/100/100 전처리 : 100/100/100	보건용마스크 기준 규격에 대한 가이드라인
분진포집효율 (파라핀오일)	개개의 측정치는 94 %이상	%	본품 : 99/99/99 전처리 : 99/99/99	보건용마스크 기준 규격에 대한 가이드라인
안면부 흡기저항	개개의 측정치는 70 Pa이하	Pa	본품 : 41/33/45 전처리 : 41/43/42	보건용마스크 기준 규격에 대한 가이드라인
누설률	시험대상자 10명의 5가지 운동결과인 총 50번 누설률 시험값 중 46번의 이상이 기준값(11.0 %) 이하이어야 한다.	%	<div> <div>본품</div> <div> 1.4/0.1/0.2/0.2/1.6 6.8/7.4/5.5/5.2/6.3 0.3/1.0/0.9/1.3/0.9 3.7/1.9/8.5/8.3/1.1 1.7/10.0/5.8/2.7/2.6 </div> </div> <div> <div>전처리</div> <div> 2.1/3.6/4.1/6.1/4.9 0.0/0.0/0.0/0.1/0.0 4.3/4.5/6.2/9.6/3.1 2.6/2.8/3.6/7.1/5.0 1.0/3.5/0.8/10.0/2.5 </div> </div>	보건용마스크 기준 규격에 대한 가이드라인
머리끈 접합부의 인장강도	검체 3개의 평균치는 10 N이상	N	50	보건용마스크 기준 규격에 대한 가이드라인

시험검사자 : 조운섭, 권수지, 권아영

기술책임자 : 왕만식

- * 위 결과는 의뢰된 시험·검사 항목만을 대상으로 한 것입니다.
- * 지면이 부족한 경우 시험·검사 항목 및 결과란은 별지로 작성 가능합니다.
- * 검사결과를 광고하거나 용기·포장 등에 표시할 때에는 시험·검사성적서 전체 내용을 모두 표시하여야 합니다.

본 시험성적서는 「식품·의약품분야 시험·검사 등에 관한 법률」에 따른 업무와 관련이 없으며, 이로 인한 법적 분쟁시 식품의약품안전처와 관련이 없음을 밝힙니다.

(사)KOTITI시험연구원



시험성적서

성적서번호 : SUQ20-00040_M1

Page 2 of 3

의뢰자	
업체명	Hunan EEXI Technology&Service Co.,Ltd.
이메일	overseas04@idore.com.cn
주 소	No.6, North of Pingtou road, Liuyang Hi-tech industrial development zone, Hunan,
전화번호	15521119635
팩스번호	

시료정보	
시료명	WK일회용 마스크
제품 번호	YX011
시험성적서 용도	안전기준준수확인용
제조자명	Hunan EEXI Technology&Service Co.,Ltd.
제조국	중국
비 고	

연령정보	
신청인 제시 사용 연령	14세 이상

시험방법	
안전기준 부속서 1-가정용 섬유제품[국가기술표준원고시 제2018-0195호(2018.06.29)]	

작성자

확인자

기술책임자



일 자 2020.05.26

일 자 2020.05.26

일 자 2020.05.26

비 고 : 1. 이 성적서는 의뢰자가 제시한 시료 및 시료명에 한정된 결과로서 전체제품에 대한 품질을 보증하지는 않습니다.

2. 이 성적서는 홍보, 선전, 광고 및 소송용으로 사용될 수 없으며, 용도 이외의 사용을 금합니다.

KCL (QINGDAO) TESTING SERVICE CO., LTD.

5TH FL, 2ND BD, NO.57, SHUYUNEAST ROAD, CHENGYANG
DISTRICT, QINGDAO, SHANDONG, CHINA 266107
Tel: (+86)532-6696-0030 Fax: (+86)532-6895-7772
E-mail: kclqd@kcl.re.kr Website: www.kcl.re.kr



시험성적서

성적서번호 : SUQ20-00041_M1

Page 2 of 3

의뢰자	
업체명	Hunan EEXI Technology&Service Co.,Ltd.
이메일	overseas04@idore.com.cn
주 소	No.6, North of Pingtuo road, Liuyang Hi-tech industrial development zone, Hunan,
전화번호	15521119635
팩스번호	

시료정보	
시료명	간편FH 일회용마스크
제품 번호	YX002
시험성적서 용도	안전기준준수확인용
제조사명	Hunan EEXI Technology&Service Co.,Ltd.
제조국	중국
비 고	

연령정보	
신청인 제시 사용 연령	14세 이상

시험방법	
안전기준 부속서 1-가정용 섬유제품[국가기술표준원고시 제2018-0195호(2018.06.29)]	

작성자

金杨英

확인자

Alex Hou

기술책임자



Lin Mingping

일 자 2020.05.26

일 자 2020.05.26

일 자 2020.05.26

- 비 고 : 1. 이 성적서는 의뢰자가 제시한 시료 및 시료명에 한정된 결과로서 전체제품에 대한 품질을 보증하지는 않습니다.
2. 이 성적서는 홍보, 선전, 광고 및 소송용으로 사용될 수 없으며, 용도 이외의 사용을 금합니다.

KCL (QINGDAO) TESTING SERVICE CO., LTD.
5TH FL, 2ND BD, NO.57, SHUYUNEAST ROAD, CHENGYANG
DISTRICT, QINGDAO, SHANDONG, CHINA 266107
Tel: (+86)532-6696-0030 Fax: (+86)532-6895-7772
E-mail: kclqd@kcl.re.kr Website: www.kcl.re.kr



시험성적서

성적서번호 : SUQ20-00042_M1

Page 2 of 3

의뢰자	
업체명	Hunan EEXI Technology&Service Co.,Ltd.
이메일	overseas04@idore.com.cn
주 소	No.6, North of Pingtou road, Liuyang Hi-tech industrial development zone, Hunan,
전화번호	15521119635
팩스번호	

시료정보	
시료명	일회용YY마스크
제품 번호	YX001
시험성적서 용도	안전기준 준수 확인용
제조자명	Hunan EEXI Technology&Service Co.,Ltd.
제조국	중국
비 고	

연령정보	
신청인 제시 사용 연령	14세 이상

시험방법	
안전기준 부속서 1-가정용 섬유제품 [국가기술표준원고시 제2018-0195호(2018.06.29)]	

작성자

金楊榮

확인자

Alex Hou

기술책임자



일 자 2020.05.26

일 자 2020.05.26

일 자 2020.05.26

- 비 고 : 1. 이 성적서는 의뢰자가 제시한 시료 및 시료명에 한정된 결과로서 전체제품에 대한 품질을 보증하지는 않습니다.
2. 이 성적서는 홍보, 선전, 광고 및 소송용으로 사용될 수 없으며, 용도 이외의 사용을 금합니다.

KCL (QINGDAO) TESTING SERVICE CO., LTD.
5TH FL, 2ND RD, NO.57, SHIYUINEAST ROAD, CHENGYANG
DISTRICT, QINGDAO, SHANDONG, CHINA 266107
Tel: (+86)532-6696-0030 Fax: (+86)532-6895-7772
E-mail: kclqd@kcl.re.kr Website: www.kcl.re.kr



시험성적서

성적서번호 : SUQ20-00043_M1

Page 2 of 5

의뢰자	
업체명	Hunan EEXI Technology&Service Co.,Ltd.
이메일	overseas04@idore.com.cn
주 소	No.6, North of Pingtuo road, Liuyang Hi-tech industrial development zone, Hunan,
전화번호	15521119635
팩스번호	

시료정보	
시료명	아동용마스크
제품 법칙	YX005
시험성적서 용도	공급자적합성확인용
제조사명	Hunan EEXI Technology&Service Co.,Ltd.
제조국	중국
비 고	

연령정보	
신청인 제시 사용 연령	6세-12세

시험방법	
공급자적합성 안전기준 부속서 15-아동용 섬유제품[산업통상자원부고시 제2018-0031호(2018.03.05)] 어린이제품 공통안전기준[산업통상자원부고시 제2017-18호(2017.01.31)]	

작성자

金杨英

확인자

Alex Hou

기술책임자



일 자 2020.05.26

일 자 2020.05.26

일 자 2020.05.26

- 비 고 : 1. 이 성적서는 의뢰자가 제시한 시료 및 시료명에 한정된 결과로서 전체제품에 대한 품질을 보증하지는 않습니다.
2. 이 성적서는 홍보, 선전, 광고 및 소송용으로 사용될 수 없으며, 용도 이외의 사용을 금합니다.
3. • 외주시험

KCL (QINGDAO) TESTING SERVICE CO., LTD.

5TH FL, 2ND BD, NO.57, SHUYUNEAST ROAD, CHENGYANG DISTRICT, QINGDAO, SHANDONG, CHINA 266107
Tel: (+86)532-6696-0030 Fax: (+86)532-6895-7772
E-mail: kclqd@kcl.re.kr Website: www.kcl.re.kr



Taiwan Ministry of Health and Welfare Medical Device Manufacturing No.

008382



西藥、醫療器材、化粧品許可證查詢

詳細處方成分 | 藥物外觀 | 仿單/外盒資料 | 授權使用 | 健保藥價查詢 | 離 開

許可證詳細內容

* * * 衛部醫器製壹字第008382號 * * *

註銷狀態		註銷日期	
註銷理由		製造許可登錄編號	
有效日期	114/06/09	發證日期	109/06/09
許可證種類	醫 器		
證字號		醫療器材級數	第一等級
通關簽審文件編號			
中文品名	"上好"醫療防護口罩(未滅菌)		
英文品名	"S. H." Medical Face Mask(Non-sterile)		
效能	限醫療器材管理辦法「醫療用衣物(I.4040)」第一等級鑑別範圍。		
醫器規格	空白		
劑型		包裝	
標籤、仿單及包裝加註			
醫器主類別一	I一般及整型外科手術裝置	醫器次類別一	I4040醫療用衣物
醫器主類別二		醫器次類別二	
醫器主類別三		醫器次類別三	
藥品分類		監視期限	
主成分略述			

<https://info.fda.gov.tw/MLMS/H0001D.aspx?Type=Lic&LicId=93008382>



Certificate

Certificate No.: USA20Q40924R0M

This is to certify that the Quality Management System of

HUNAN EEXI TECHNOLOGY & SERVICE CO., LTD.

Unified Social Credit Code: 91430181MA4PHUE510

Registered/Office Address: No.6 North of Pingtuo Road, Liuyang Hi-Tech Industrial Development Zone, Hunan, China

Production Address: Floor 1 and 2, Building 3, No.6 North of Pingtuo Road, Liuyang Hi-Tech Industrial Development Zone, Hunan, China

Has been audited to conform to the following Quality Management System standard

ISO 9001:2015

This Quality Management System is valid for the

The production of disposable medical masks (non-sterile), disposable medical masks (sterile), medical surgical masks (non-sterile) and medical surgical masks (sterile) (within the company's license scope)

Date of initial issuance: Apr. 20, 2020

Date of issuance: Apr. 20, 2020

Date of renewal: May. 12, 2020

Date of expiry: Apr. 19, 2023

Issued by: *Wu Fengru*

Beijing East Allreach Certification Center Co., Ltd.



The certificate will remain valid only if the certified organization accepts surveillance audit at regular intervals and is audited to be qualified. The information of this certificate is available at EACC website (www.eacc.com.cn) and CNCA's official website (www.cnca.gov.cn), and it's also available by scanning the QR Code in the lower right corner.

EACC address: 1st Floor, No. 121 building, No. 17, Jingshengniansi Street, Jinqiao Science & Technology Industrial Base, Tongzhou Park of Zhongguancun Science & Technology Zone, Tongzhou District, Beijing 101102





Certificate

Certificate No.: USA20E40925R0M

This is to certify that the Environmental Management System of

HUNAN EEXI TECHNOLOGY & SERVICE CO., LTD.

Unified Social Credit Code: 91430181MA4PHUE510

Registered/Office Address: No.6 North of Pingtuo Road, Liuyang Hi-Tech Industrial
Development Zone, Hunan, China

Production Address: Floor 1 and 2, Building 3, No.6 North of Pingtuo Road, Liuyang Hi-Tech
Industrial Development Zone, Hunan, China

Has been audited to conform to the following Environmental
Management System standard

ISO 14001:2015

This Environmental Management System is valid for the
The production of disposable medical masks (non-sterile), disposable medical
masks (sterile), medical surgical masks (non-sterile) and medical surgical
masks (sterile) (within the company's license scope) and related
environmental management activities of involved sites

Date of initial issuance: Apr. 20, 2020

Date of issuance: Apr. 20, 2020

Date of renewal: May. 12, 2020

Date of expiry: Apr. 19, 2023

Issued by: *Wu Fanyue*

Beijing East Allreach Certification Center Co., Ltd.



The certificate will remain valid only if the certified organization accepts surveillance audit at regular intervals and is audited to be qualified. The information of this certificate is available at EACC website (www.eacc.com.cn) and CNCA's official website (www.cnca.gov.cn), and it's also available by scanning the QR Code in the lower right corner.

EACC address: 1st Floor, No. 121 building, No. 17, Jingshengnansi Street, Jinqiao Science & Technology Industrial Base, Tongzhou Park of Zhongguancun Science & Technology Zone, Tongzhou District, Beijing 101102



Certificate CN20/42091

The management system of

Hunan EEXI Technology & Service Co., Ltd.

No. 6, North of Pingtuo Road, Liuyang Hi-tech Industrial Development Zone,
Hunan Province, 410300, P.R. China

has been assessed and certified as meeting the requirements of

ISO 13485:2016 EN ISO 13485:2016

For the following activities

**Design and manufacture of non-sterile surgical face mask,
non-sterile single use medical face mask and
non-sterile medical protective face mask**

This certificate is valid from 16 June 2020 until 15 June 2023
and remains valid subject to satisfactory surveillance audits.

Re certification audit due before 5 June 2023

Issue 1. Certified since 16 June 2020

Authorised by



SGS United Kingdom Ltd
Rossmore Business Park, Ellesmere Port, Cheshire, CH65 3EN, UK
t +44 (0)151 350-6666 f +44 (0)151 350-6600 www.sgs.com

HC SGS 13485 2016 0118

Page 1 of 1





Sponsor:
Eddie Yam
Intertek Testing Services Hong Kong Ltd.
1/F, Garment Centre, 576 Castle Peak Road
Kowloon,
HONG KONG

Viral Filtration Efficiency (VFE) Final Report

Test Article: modified non-woven
colour: White
Style #1001
Study Number: 1280865-S01
Study Received Date: 25 Mar 2020
Testing Facility: Nelson Laboratories, LLC
6280 S. Redwood Rd.
Salt Lake City, UT 84123 U.S.A.
Test Procedure(s): Standard Test Protocol (STP) Number: STP0007 Rev 16
Deviation(s): None

Summary: The VFE test is performed to determine the filtration efficiency of test articles by comparing the viral control counts upstream of the test article to the counts downstream. A suspension of bacteriophage Φ X174 was aerosolized using a nebulizer and delivered to the test article at a constant flow rate and fixed air pressure. The challenge delivery was maintained at $1.1 - 3.3 \times 10^3$ plaque forming units (PFU) with a mean particle size (MPS) of $3.0 \mu\text{m} \pm 0.3 \mu\text{m}$. The aerosol droplets were drawn through a six-stage, viable particle, Andersen sampler for collection. The VFE test procedure was adapted from ASTM F2101.

All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Test Side: Either
Test Area: $\sim 40 \text{ cm}^2$
VFE Flow Rate: 28.3 Liters per minute (L/min)
Conditioning Parameters: $85 \pm 5\%$ relative humidity (RH) and $21 \pm 5^\circ\text{C}$ for a minimum of 4 hours
Positive Control Average: 1.6×10^3 PFU
Negative Monitor Count: <1 PFU
MPS: $2.9 \mu\text{m}$

Study Director

James W. Luskin

Study Completion Date



1280865-S01

801-290-7500 | nelsonlabs.com | sales@nelsonlabs.com

myf

FRTD007-0001 Rev 16

Page 1 of 2

Results:

Test Article Number	Percent VFE (%)
1	>99.9 ^a
2	>99.9 ^a
3	>99.9 ^a
4	>99.9 ^a
5	>99.9 ^a

^a There were no detected plaques on any of the Andersen sampler plates for this test article.

The filtration efficiency percentages were calculated using the following equation:

$$\% VFE = \frac{C - T}{C} \times 100$$

C = Positive control average

T = Plate count total recovered downstream of the test article

Note: The plate count total is available upon request

Bacterial Filtration Efficiency with Increased Delivery Challenge (BFE) in ASTM F2101 and EN14683
Proven that Curie technology can effectively filter increased challenge of bacteria (99.8%)



Sponsor:
Wong Chak Ming
Hong Kong Medical Supply Limited
Unit A3, 2F Mai Wah Industrial Bldg.
1-7 Wah Sing Street
Hong Kong,
CHINA

Bacterial Filtration Efficiency (BFE) Final Report

Test Article: HKMSLMASK000
Purchase Order: HKMSLPO20200326
Study Number: 1282265-S01
Study Received Date: 28 Mar 2020
Testing Facility: Nelson Laboratories, LLC
6280 S. Redwood Rd.
Salt Lake City, UT 84123 U.S.A.
Test Procedure(s): Standard Test Protocol (STP) Number: STP0004 Rev 18
Deviation(s): None

Summary: The BFE test is performed to determine the filtration efficiency of test articles by comparing the bacterial control counts upstream of the test article to the bacterial counts downstream. A suspension of *Staphylococcus aureus* was aerosolized using a nebulizer and delivered to the test article at a constant flow rate and fixed air pressure. The challenge delivery was maintained at 3.5×10^3 colony forming units (CFU) with a mean particle size (MPS) of $3.0 \pm 0.3 \mu\text{m}$. The aerosols were drawn through a six-stage, viable particle, Andersen sampler for collection. This test method complies with ASTM F2101-19 and EN 14683:2019, Annex B; with the exception of the **higher challenge level**, which may represent a **more severe test**.

All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Test Side: Inside
BFE Test Area: $\sim 40 \text{ cm}^2$
BFE Flow Rate: 28.3 Liters per minute (L/min)
Conditioning Parameters: $85 \pm 5\%$ relative humidity (RH) and $21 \pm 5^\circ\text{C}$ for a minimum of 4 hours
Test Article Dimensions: $\sim 176 \text{ mm} \times \sim 160 \text{ mm}$
Positive Control Average: $3.5 \times 10^3 \text{ CFU}$
Negative Monitor Count: $< 1 \text{ CFU}$
MPS: $3.0 \mu\text{m}$

The positive control average was out of specification per STP0004 Rev 18 section 6.1 which states, "The BFE positive control average shall be maintained at $1.7\text{-}3.0 \times 10^3 \text{ CFU}$." Testing with a **more severe challenge** to the test articles represents a worse case. The sponsor accepted the use of the **higher challenge**; therefore, the results are considered valid at the testing conditions that occurred.



Alissa Sanders
Study Director

for
James W. Luskin

20 Apr 2020
Study Completion Date



1282265-S01

801-290-7500 | nelsonlabs.com | sales@nelsonlabs.com

brd

FRT0004-0001 Rev 22
Page 1 of 2

Results:

Test Article Number	Percent BFE (%)
1	99.8
2	99.8
3	99.8
4	99.8
5	99.8

The filtration efficiency percentages were calculated using the following equation:

$$\% BFE = \frac{C - T}{C} \times 100$$

C = Positive control average

T = Plate count total recovered downstream of the test article

Note: The plate count total is available upon request

Viral Filtration Efficiency (VFE) in ASTM F2101

Proven that Curie technology can effectively **filter virus (>99.9a%)**



Number: HKGT05112613-S1

TEST REPORT

Applicant: CURIE LIMITED
B3-1 G/F
SUPERLUCK INDL CTR PHASE 2
57 SHA TSUI RD
TSUEN WAN NT HK

Date: Apr 22, 2020
This is to supersede report no.
HKGT05112613 dated Apr 21,
2020

Attn: ALDRIN OR

Sample Description As Declared :

No. Of Sample : Several
Buyer's Name : -
Agent's Name : -
Manufacturer's Name : Curie Limited
Sample Description : Curie Ultrahigh-Efficiency Viral Filter超高效病毒濾材
Colour : White
Style No. : 1001
Order No. / PO No. : -
Product End Uses : -
Fibre Content : Nonwoven
Fabric/GMT Weight : 20g
Ref. : -

Date Received/Date Test Started : Apr 15, 2020

Applicant's Provided Care Instruction/Label :
-



Number: HKGT05112613-S1

TEST REPORT

Original Sample Photo:



For any queries on this report, you are welcome to contact our customer service representatives:

US3

Angie Yu (852) 98639123 or email to angie.yu@intertek.com

TEST REPORT

Tests Conducted (As Requested By The Applicant)

1 Evaluation of Viral Filtration Efficiency (VFE):

Summary: The VFE test is performed to determine the filtration efficiency of test articles by comparing the viral control counts upstream of the test article to the counts downstream. A suspension of bacteriophage ΦX174 was aerosolized using a nebulizer and delivered to the test article at a constant flow rate and fixed air pressure. The challenge delivery was maintained at $1.1 - 3.3 \times 10^3$ plaque forming units (PFU) with a mean particle size (MPS) of $3.0 \pm 0.3 \mu\text{m}$. The aerosols droplets were drawn through a six-stage, viable particle, Andersen sampler for collection. The VFE test procedure was adapted from ASTM F2101.

All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Test side: Either

Test Area: $\sim 40 \text{ cm}^2$

VFE Flow Rate: 28.3 Liters per minute (L/min)

Conditioning Parameters: $85 \pm 5\%$ relative humidity (RH) and $21 \pm 5^\circ\text{C}$ for a minimum of 4 hours

Positive Control Average: 1.6×10^3 PFU

Negative Monitor Count: <1 PFU

MPS: $2.9 \mu\text{m}$

TEST REPORT

Tests Conducted (As Requested By The Applicant)

Evaluation of Viral Filtration Efficiency (Cont'd)

Result:

Test Article Number	Percent VFE (%)
1	$>99.9^a$
2	$>99.9^a$
3	$>99.9^a$
4	$>99.9^a$
5	$>99.9^a$

^a There were no detected plaques on any of the Andersen sampler plates for this test article.

The filtration efficiency percentages were calculated using the following equation:

$$\%VFE = \frac{C - T}{C} \times 100$$

C= Positive control average

T= Plate count total recovered downstream of the test article

Note: The plate count total is available upon request

Remark: The test was conducted by competent subcontractor lab.

Bacterial Filtration Efficiency (BFE) in ASTM F2101 and EN14683

Proven that Curie technology can effectively filter bacteria (99.9%), complying ASTM F2101 Level 3 and EN14683 Type IIR

Air Exchange Pressure in ASTM F2101 / EN14683

Proven that air exchange pressure of KV99 masks can comply with standards of ASTM F2101 Level 3 and EN14683 Type IIR



Sponsor:
Wing Wah Medicine Group Holdings Limited
1/F Ching Cheong Industrial Building
1-7 Kwai Chong Rd
Kwai Chung, NT

Bacterial Filtration Efficiency (BFE) and Differential Pressure (Delta P) Final Report

Test Article: 4 Ply Disposable Medical Mask
Colour: White
Fiber Content: Curie Biohazard Filter + SS Non-Woven
Purchase Order: R904838210
Study Number: 1334934-S01
Study Received Date: 25 Aug 2020
Testing Facility: Nelson Laboratories, LLC
6280 S. Redwood Rd.
Salt Lake City, UT 84123 U.S.A.
Test Procedure(s): Standard Test Protocol (STP) Number: STP0004 Rev 18
Deviation(s): None

Summary: The BFE test is performed to determine the filtration efficiency of test articles by comparing the bacterial control counts upstream of the test article to the bacterial counts downstream. A suspension of *Staphylococcus aureus* was aerosolized using a nebulizer and delivered to the test article at a constant flow rate and fixed air pressure. The challenge delivery was maintained at $1.7 - 3.0 \times 10^3$ colony forming units (CFU) with a mean particle size (MPS) of $3.0 \pm 0.3 \mu\text{m}$. The aerosols were drawn through a six-stage, viable particle, Andersen sampler for collection. This test method complies with ASTM F2101-19 and EN 14683:2019, Annex B.

The Delta P test is performed to determine the breathability of test articles by measuring the differential air pressure on either side of the test article using a manometer, at a constant flow rate. The Delta P test complies with EN 14683:2019, Annex C and ASTM F2100-19.

All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Test Side: Inside
BFE Test Area: $\sim 40 \text{ cm}^2$
BFE Flow Rate: 28.3 Liters per minute (L/min)
Delta P Flow Rate: 8 L/min
Conditioning Parameters: $85 \pm 5\%$ relative humidity (RH) and $21 \pm 5^\circ\text{C}$ for a minimum of 4 hours
Test Article Dimensions: $\sim 175 \text{ mm} \times \sim 165 \text{ mm}$
Positive Control Average: 2.0×10^3 CFU
Negative Monitor Count: <1 CFU
MPS: $3.0 \mu\text{m}$



Leah Tiberius electronically approved for
Study Director

James Luskin

18 Sep 2020 16:53 (+00:00)
Study Completion Date and Time

Results:

Test Article Number	Percent BFE (%)
1	>99.9
2	99.9
3	>99.9 ^a
4	>99.9
5	99.8

^a There were no detected colonies on any of the Andersen sampler plates for this test article.

Test Article Number	Delta P (mm H ₂ O/cm ²)	Delta P (Pa/cm ²)
1	5.2	50.6
2	5.2	50.5
3	5.4	52.6
4	5.3	51.5
5	4.8	47.5

The filtration efficiency percentages were calculated using the following equation:

$$\% BFE = \frac{C - T}{C} \times 100$$

C = Positive control average
 T = Plate count total recovered downstream of the test article
 Note: The plate count total is available upon request

Microbial Cleanliness (Bioburden) of Medical Masks in EN14683

Proven that total bioburden of KV99 can comply with standard of EN14683 Type IIR (<6 cfu)



Sponsor:
Wing Wah Medicine Group Holdings Limited
1/F Ching Cheong Industrial Building
1-7 Kwai Chong Rd
Kwai Chung, NT
HONG KONG

Microbial Cleanliness (Bioburden) of Medical Masks Final Report

Test Article: 4 Ply Disposable Medical Mask
Colour: White
Fiber Content: Curie Biohazard Filter + SS Non-Woven
Purchase Order: R904838210
Study Number: 1334930-S01
Study Received Date: 25 Aug 2020
Testing Facility: Nelson Laboratories, LLC
6280 S. Redwood Rd.
Salt Lake City, UT 84123 U.S.A.
Test Procedure(s): Standard Test Protocol (STP) Number: STP0036 Rev 15
Customer Specification Sheet (CSS) Number: 202002096 Rev 01
Deviation(s): None

Summary: The testing was conducted in accordance with EN 14683:2019, with the exception of approximate volumes of eluent used when performing the extraction procedure and a temperature range of 30-35°C used for aerobic incubation.

When bioburden results are calculated using a software program, manual calculations may differ slightly due to rounding. The counts determined on products are colony forming units and may not always reflect individual microorganisms. The sponsor performs any statistical analysis and determines the acceptable limits. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.



Robert Putnam electronically approved
Study Director

Robert Putnam

24 Sep 2020 02:10 (+00:00)
Study Completion Date and Time

Results:

Unit Number	Weight (g)	Aerobic	Fungal	Total Bioburden (CFU/mask)	Total Bioburden (CFU/g)
1	4.3	<3	<3	<5.7	<1.3
2	4.3	6	<3	<8.9	<2.1
3	4.2	<3	<3	<5.8	<1.4
4	4.2	3	<3	<6.0	<1.4
5	4.2	<3	<3	<6.0	<1.4
Recovery Efficiency	UTD ^a				

< = No Organisms Detected

UTD = Unable to Determine

Note: The results are reported as colony forming units per test article.

^a UTD due to zero count on the first rinse. An alternative method or inoculated product recovery efficiency is recommended.

Method Suitability:

Organism	Percentage
<i>Bacillus atrophaeus</i>	0%

Test Method Acceptance Criteria: If applicable, anaerobic controls are acceptable for the bioburden test results. The number of masks to be tested shall be a minimum of 5 or more to meet an acceptable quality level of 4%. The bioburden of the medical mask shall be < 30 CFU/g tested.

Procedure:

Positive Controls/Monitors: *Bacillus atrophaeus*
 Extract Fluid: Peptone Tween®
 Extract Fluid Volume: ~300 mL
 Extract Method: Orbital Shaking for 15 minutes at 250 rpm
 Plating Method: Membrane Filtration
 Agar Medium: Tryptic Soy Agar
 Potato Dextrose Agar
 Recovery Efficiency: Exhaustive Rinse Method
 Aerobic Bacteria: Plates were incubated 3-7 days at 30-35°C, then enumerated.
 Fungal: Plates were incubated 5-7 days at 20-25°C, then enumerated.

Synthetic Blood Penetration Pressure in ASTM F2101 / ASTM F1862 / EN14683

Proven that KV99 mask can effectively intercept fluid (160mmHg) from penetration, complying ASTM F2101 Level 3 and EN14683 Type IIR



Sponsor:
Wing Wah Medicine Group Holdings Limited
1/F Ching Cheong Industrial Building
1-7 Kwai Cheong Road
Kwai Chung, NT

Synthetic Blood Penetration Resistance Final Report

Test Article: 4 Ply Disposable Medical Mask
Colour: White
Fiber Content: Curie Biohazard Filter + SS Non-Woven
Purchase Order: R904838210
Study Number: 1334931-S01
Study Received Date: 25 Aug 2020
Testing Facility: Nelson Laboratories, LLC
6280 S. Redwood Rd.
Salt Lake City, UT 84123 U.S.A.
Test Procedure(s): Standard Test Protocol (STP) Number: STP0012 Rev 09
Deviation(s): None

Summary: This procedure was performed to evaluate surgical facemasks and other types of protective clothing materials designed to protect against fluid penetration. The purpose of this procedure is to simulate an arterial spray and evaluate the effectiveness of the test article in protecting the user from possible exposure to blood and other body fluids. The distance from the target area surface to the tip of the cannula is 30.5 cm. A test volume of 2 mL of synthetic blood was employed using the targeting plate method.

This test method was designed to comply with ASTM F1862 and ISO 22609 (as referenced in EN 14683:2019 and AS4381:2015) with the following exception: ISO 22609 requires testing to be performed in an environment with a temperature of $21 \pm 5^{\circ}\text{C}$ and a relative humidity of $85 \pm 10\%$. Instead, testing was performed at ambient conditions within one minute of removal from the environmental chamber held at those parameters.

All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Number of Test Articles Tested: 32
Number of Test Articles Passed: 32
Test Side: Outside
Pre-Conditioning: Minimum of 4 hours at $21 \pm 5^{\circ}\text{C}$ and $85 \pm 5\%$ relative humidity (RH)
Test Conditions: 23.2°C and 22% RH

Results: Per ASTM F1862 and ISO 22609, an acceptable quality limit of 4.0% is met for a normal single sampling plan when ≥ 29 of 32 test articles show passing results.

Test Pressure: 160 mmHg (21.3 kPa)

Test Article Number	Synthetic Blood Penetration
1-32	None Seen

Study Director

For
James W. Luskin

17 SEP 2020
Study Completion Date



1334931-S01

801-290-7500 | nelsonlabs.com | sales@nelsonlabs.com

brd

FRT0012-0002 Rev 13
Page 1 of 1

Sub-Micron Particulate Filtration Efficiency (0.1µm PSL) in

ASTM F2101 / ASTM F2299 / EN14683

Proven that KV99 masks can effectively filter 0.1µm sub-micron particulate (>99.8%), complying ASTM F2101 Level 3 and EN14683 Type IIR



Sponsor:
Wing Wah Medicine Group Holdings Limited
1/F Ching Cheong Industrial Building,
1-7 Kwai Chong Rd,
Kwai Chung, NT

Latex Particle Challenge Final Report

Test Article: 4 Ply Disposable Medical Mask
Colour: White
Fiber Content: Curie Biohazard Filter + SS Non-Woven
Purchase Order: R904838210
Study Number: 1334932-S01
Study Received Date: 25 Aug 2020
Testing Facility: Nelson Laboratories, LLC
6280 S. Redwood Rd.
Salt Lake City, UT 84123 U.S.A.
Test Procedure(s): Standard Test Protocol (STP) Number: STP0005 Rev 08
Deviation(s): None

Summary: This procedure was performed to evaluate the non-viable particle filtration efficiency (PFE) of the test article. Monodispersed polystyrene latex spheres (PSL) were nebulized (atomized), dried, and passed through the test article. The particles that passed through the test article were enumerated using a laser particle counter.

A one-minute count was performed, with the test article in the system. A one-minute control count was performed, without a test article in the system, before and after each test article. Control counts were performed to determine the average number of particles delivered to the test article. The filtration efficiency was calculated using the number of particles penetrating the test article compared to the average of the control values. During testing and controls, the air flow rate is maintained at 1 cubic foot per minute (CFM) \pm 5%.

The procedure employed the basic particle filtration method described in ASTM F2299, with some exceptions; notably the procedure incorporated a non-neutralized challenge. In real use, particles carry a charge, thus this challenge represents a more natural state. The non-neutralized aerosol is also specified in the FDA guidance document on surgical face masks. All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Test Side: Inside
Area Tested: 91.5 cm²
Particle Size: 0.1 µm
Laboratory Conditions: 21°C, 28% relative humidity (RH) at 1813; 21°C, 28% RH at 1911
Average Filtration Efficiency: 99.86%
Standard Deviation: 0.038



Trang Truong electronically approved for
Study Director

Curtis Gerow

16 Sep 2020 22:26 (+00:00)
Study Completion Date and Time

Results:

Test Article Number	Test Article Counts	Average Control Counts	Filtration Efficiency (%)
1	9	10,448	99.914
2	17	11,356	99.85
3	22	12,202	99.82
4	19	12,264	99.85
5	13	12,067	99.89



Issuing Laboratory:

Intertek Testing Services Hong Kong Ltd.

Hong Kong Accreditation Service (HKAS) has accredited this laboratory (Reg. No. HOKLAS 005) under Hong Kong Laboratory Accreditation Scheme (HOKLAS) for specific laboratory activities as listed in the HOKLAS Directory of Accredited Laboratories.



Number: HKGT05148314

TEST REPORT

Applicant: WING WAH MEDICINE GROUP HOLDINGS LIMITED
1/F CHING CHEONG IND BLDG
1-7 KWAI CHEONG RD
KWAI CHUNG NT
HK

Date: Sep 28, 2020

Attn: ANNIE AU

Sample Description As Declared :

No. Of Sample : Several
Buyer's Name : -
Agent's Name : -
Manufacturer's Name : -
Sample Description : 4 Ply Disposable Medical Mask
Colour : White
Style No. : -
Order No. / PO No. : -
Product End Uses : -
Fibre Content : Curie Biohazard Filter + SS Non-Woven
Fabric/GMT Weight : -
Ref. : -

Date Received/Date Test Started : Aug 18, 2020

Applicant's Provided Care Instruction/Label :

For and on behalf of
Intertek Testing Services HK Ltd


WONG Sai Ming
Technical Supervisor II



Page 1 Of 3



Issuing Laboratory:

Intertek Testing Services Hong Kong Ltd.

Hong Kong Accreditation Service (HKAS) has accredited this laboratory (Reg. No. HOKLAS 005) under Hong Kong Laboratory Accreditation Scheme (HOKLAS) for specific laboratory activities as listed in the HOKLAS Directory of Accredited Laboratories.



Number: HKGT05148314

TEST REPORT

Tests Conducted (As Requested By The Applicant)

- 1 Flammability of 45 degree angle test (US 16 CFR Part 1610 - Standard for the Flammability of Clothing Textiles):

☒ Plain Surface ☐ Raised Surface

Prelim Plain Surface:
Length : IBE
Width : IBE

Burn Direction:	<input checked="" type="checkbox"/> Length <input type="checkbox"/> Width
	Original (seconds)
1.	IBE
2.	IBE
3.	DNI
4.	IBE
5.	DNI
6.	-
7.	-
8.	-
9.	-
10.	-



Issuing Laboratory:

Intertek Testing Services Hong Kong Ltd.

Hong Kong Accreditation Service (HKAS) has accredited this laboratory (Reg. No. HOKLAS 005) under Hong Kong Laboratory Accreditation Scheme (HOKLAS) for specific laboratory activities as listed in the HOKLAS Directory of Accredited Laboratories.



Number: HKGT05148314

TEST REPORT

Tests Conducted (As Requested By The Applicant)

Flammability of 45 degree angle test (Cont'd)

Classification: ☒ Class 1, Normal Flammability
☐ Class 2, Intermediate Flammability, Raised Surface
☐ Class 3, Rapid And Intense Burning,

Explanation Of Flammability Results:

DNI Did not ignite.
IBE Ignited but extinguished.
___ sec. Actual burn time measured and recorded by the timing device.
SF uc Surface flash, under the stop thread, but does not break the stop thread.
SF pw Surface flash, part way. No time shown because the surface flash did not reach the stop thread.
SF poi Surface flash, at point of impingement only (equivalent to "Did not ignite" for plain surfaces).
___ SF only Time in seconds, surface flash only. No damage to the base fabric.
___ SFBB Time in seconds, surface flash base burn starting at places other than the point of impingement as a result of surface flash
___ SFBB poi Time in seconds, surface flash base burn starting at the point of impingement.
___ SFBB poi* Time in seconds, surface flash base burn possibly starting at the point of impingement. The asterisk is accompanied by the following statement: "Unable to make absolute determination as to source of base burns." This statement is added to the result of any specimen if there is a question as to origin of the base burn.

End of Report

When a statement of conformity to a specification or standard is provided on test report, the decision rule shall be applied. For details, please refer to the latest version of Intertek's "Decision Rule Information" and is available on Intertek's website: <https://intertekhk.org/by/decision-rules-info>. If decision rule already inhered in the requested specification or standard, Intertek's "Decision Rule Information" is not applicable.

This report is for the exclusive use of Intertek's Client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to and subject to our standard Terms and Conditions which can be obtained at our website: <http://www.intertek.com/terms/>. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this report. Intertek is responsible for all the information provided in the reports, except when information is provided by the Client or when the Client requires the item to be tested acknowledging a deviation from specified conditions that can affect the validity of results.

The observations and test results in this report are relevant to the sample(s) tested and submitted by client. The report is not intended to be a recommendation for any particular course of action, you are responsible for acting as you see fit on the basis of the report results. This report does not discharge or release you from your legal obligations and duties to any other person. Only the Client is authorized to permit copying or distribution of this report and the report shall not be reproduced except in full. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. This report by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program.






Report No: ATCCR20081010F

Test Report

Sample Category	Curie Ultrahigh-Efficiency Viral Filter for KV-99
Client	Curie Limited
Test Category	Test Entrust
Date of Report	2020.08.18

Report No: ATCCR20081010F

Detection Information

Client	Curie Limited		Sample Source	Inspect
Client address	Room C,23/F,Tsuen Tung Factory Building,38-40 Chai Wan KOK Street, Tsuen Wan,Hong Kong		Sample State	Normal
Date of Receives samples	2020.08.10		Date(s) of tests	2020.08.10-2020.08.18
Sample No	ATCCR20081010F-0810CP01			
Category	Test Project	Test Standard and Method		Test Instruments
Curie Ultrahigh-Efficiency Viral Filter for KV-99	Antiviral Activity Value (COVID-19)	ISO 18184:2019 Textiles Determination of antiviral		Biosafety Cabinet
End				
Remarks	Production units: Curie Limited Trademarks: Curie Date of production: 2020.06.09 Sample model: Curie Ultrahigh-Efficiency Viral Filter for KV-99 Sample batch: 1001			
Report Preparer: 刘畅	Authorized Signatory: 刘畅			
Report Reviewer: 李红	t Date of Issues Report: 2020.08.18			
(Special Chapter for Inspection and Inspection)				

Test results

Virus Types	(NO)	$\lg(V_{a0h})$ ($\lg\text{TCID}_{50}/\text{mL}$)	$\lg(V_{b2h})$ ($\lg\text{TCID}_{50}/\text{mL}$)	$\lg(V_{c2h})$ ($\lg\text{TCID}_{50}/\text{mL}$)
COVID-19 virus MDCK cells	1	6.73	6.68	3.7
	2	6.68	6.56	4
	3	6.7	6.57	3.9
Average Value of $\lg\text{TCID}_{50}/\text{mL}$		6.70	6.61	3.88
Antiviral Activity Value		2.72		
Antiviral Activity Rate (%)		99.81		



End



Beijing Shantong Medical Testing Laboratory

Declaration of Test Results

Beijing Shantong Medical Testing Lab "BSMTL"
hereby declares that the test item described below
has been tested by BSMTL and complies with the
requirements of

ISO 18184: 2019 Textile Determination of Antiviral

The complete detail of the tests performed and the
results are recorded in

Report No: ATCCR20081010F Dated: 18.08.2020

Description of item tested: Curie Ultrahigh-Efficiency Viral Filter for KV-99

Virus Tested: SARS-COV-2 / COVID-19 MDCK Cells

Summary of test results -

Antiviral Activity Value: 2.72

Antiviral Activity Rate: 99.81%

Submitted by: Curie Limited
Room C, 23/F, Tsuen Tung Factory
Building, 38-40 Chai Wan Kok Street,
Tsuen Wan, Hong Kong SAR

Declaration authorised by:

Name:

赵志同

Title:

实验室负责人

Date: 02/09/2020



Attention is drawn to the conditions upon which this declaration
is issued, namely:

1. This declaration does not indicate provide or imply any
measure of Approval, Certification, Supervision, Control or
Surveillance by BSMTL to this or any related product.

2. This Declaration applied only to the particular sample tested
and to the specific tests carried out as detailed in the Report
referred to above.

3. The general and specific conditions of the BSMTL
Conditions of Contract for Testing, apply in all respects.

Beijing Shantong Medical Testing Laboratory Co. Ltd., Fangshan, Beijing, China



统一社会信用代码
91110111MA01A4KK4D

营业执照

(副本) (1-1)



扫描二维码登录
“国家企业信用
信息公示系统”
了解更多信息、
备案、许可、监
管信息

名称 北京普通医学检验实验室有限公司
类型 有限责任公司(法人独资)
法定代表人 杨益
经营范围 医学检验医疗服务；技术开发、技术转让、技术咨询、技
术服务。（企业依法自主选择经营项目，开展经营活动；
医疗服务以及依法须经批准的项目，经相关部门批准后依
批准的内容开展经营活动；不得从事本市产业政策禁止和
限制类项目的经营活动。）

注册资本 600万元
成立日期 2018年01月24日
营业期限 2018年01月24日至2048年01月23日
住所 北京市房山区拱辰街道办事处学园北街11号综合
服务楼一层5106

登记机关



2019年05月28日

国家企业信用信息公示系统网址：<http://www.gsxt.gov.cn>

市场主体应当于每年1月1日至6月30日通过
国家企业信用信息公示系统报送公示年度报告。

国家市场监督管理总局监制



中华人民共和国

医疗机构执业许可证

机构名称 北京善通医学检验实验室

法定代表人 杨益

地址 北京市房山区拱辰街道办事处学院北街11号综合服务楼一层

主要负责人 赵志国

诊疗科目 医学检验科;临床免疫、血清学专业;
临床细胞分子遗传学专业*****

登记号 007913110111417919

有效期限 自 2019年 07月 02日至 2022年 12月 30日

该医疗机构经核准登记, 准予执业

中华人民共和国国家卫生健康委员会



发证机关 北京市房山区卫生健康委员会

发证日期 2019年 8月 14日



开户许可证

核准号: J1000210464502

编号: 1000-03577720

经审核, 北京善通医学检验实验室有限公司

符合开户条件, 准予

开立基本存款账户。

法定代表人(单位负责人) 杨益

开户银行 兴业银行股份有限公司北京房山支行

账号 321580100100032556

发证机关(盖章)

2018年11月16日



北京市卫生健康委员会

北京市卫生健康委员会关于同意

密云区医院等 11 家检测机构

开展新型冠状病毒核酸检测的通知

西城区、朝阳区、丰台区、房山区、顺义区、大兴区、密云区卫生健康委，经济开发区，市疾控中心，市医学检验质控中心，各相关医疗机构：

根据北京市密云区医院、北京市西城区展览路医院、北京市丰台区铁营医院、北京市朝阳区三环肿瘤医院、北京朝阳急诊抢救中心、北京朝阳中西医结合急诊抢救中心、北京市大兴区中西医结合医院、北京北亚骨科医院、北京德威铭达医学检验所、**北京善通医学检验实验室**和北京索真医学检验实验室等 11 家检测机构（以下简称 11 家检测机构）提交的开展新冠病毒核酸检测的申请，结合专家评估意见，经研究，现就有关事项通知如下：

一、同意 11 家检测机构开展新型冠状病毒核酸检测工作。

二、11 家检测机构要严格按照国家和本市关于开展新型冠状病毒核酸检测、生物安全防护、生物样本资源管理的有

北京市病原微生物实验室及实验室活动备案通知书

京房山卫实验室备字[2020]第043号

北京善通医学检验实验室有限公司：

你单位于2020年06月27日提交的北京市病原微生物实验室及实验活动备案材料如下：

1. ☒ 《北京市病原微生物实验室及实验室活动备案表》；
2. ☒ 实验室或实验室设立单位的法人资格证明；
3. ☒ 实验室设立单位生物安全组织管理框架图；
4. ☒ 实验室布局平面图；

经本机关审查，认为申请材料齐全、符合《北京市病原微生物实验室及实验活动备案管理办法》的要求，决定予以备案。

卫生健康行政部门（印章）

2020年06月27日

备注：此备案旨在了解你单位实验室及其实验活动基本状况，不作为审批依据。请你单位备案后，严格按照《中华人民共和国传染病法》、《病原微生物实验室生物安全管理条例》和《人间传染的高致病性病原微生物实验室和实验活动生物安全审批管理办法》等相关法律法规规定，从事相关实验活动，规范实验室管理。

北京市卫生健康委员会制定

Sub-Micron Particulate Filtration Efficiency (0.1µm PSL) in

ASTM F2101 / ASTM F2299 / EN14683

Proven that KV99 masks can effectively filter 0.1µm sub-micron particulate (>99%), complying ASTM F2101 Level 3 and EN14683 Type IIR

Air Exchange Pressure in ASTM F2101 / EN14683

Proven that air exchange pressure of KV99 masks can comply with standards of ASTM F2101 Level 3 and EN14683 Type IIR



TEST REPORT TUCHENG

Date: Aug.31,2020 Date of Receipt: Aug.24,2020

Report No.: TFF9H598 Quantity: 1PC Page Order/Pages: (P1/5) Ref. No.: NIL

Report Title: Curie Limited(U3104) Item: Mask

Address: No. 108, Ganhe Rd., Xitun Dist., Taichung City 407, Taiwan

Test Items		Test Results	Test Methods
Sub-Micron Particulate Filtration Efficiency(%) (0.1µm PSL)	1	99.49	ASTM F2100-2019 9.3
	2	99.45	ASTM F2299-2017
	3	99.46	Flow rate:28.1
	4	99.48	(Liter/min)
	5	99.48	
	Ave.	99.47	
Air Exchange Pressure (mmH ₂ O/cm ²)	1	5.3	ASTM F2100-2019 9.2
	2	5.3	EN 14683:2019 Annex C
	3	4.9	
	4	5.1	
	5	5.3	
Flammability (as Received)		DNI	ASTM F2100-2019 9.5 CPSC 16 CFR 1610-2008

Note: 1mmH₂O=9.8Pa.

Note: Air Exchange Pressure takes 5 pieces for testing.

Note: Flammability takes 20 samples for testing.

Note: "DNI":Did Not Ignite.

Note: Sample description is given by the client: Brandname: Curie

Model: Curie KV99 Face Mask

Hong Kong address: Unit C, 23F, Tsun Tung Factory Bldg 38-40

Chai Wan Kok St, Tsuen Wan. NT Hong Kong

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.

- Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.
- 2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.
- 3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of
Taiwan Textile Research Institute

Jui-hung Kao
Director,
Department of Testing and
Certification

Department of Testing and Certification, Taiwan Textile Research Institute
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)
Tel : +886-2-22670321 ext. 7107, 7110
Fax : +886-2-22675108 , +886-2-22689839



TEST REPORT TUCHENG

Date: Aug. 31, 2020 Date of Receipt: Aug. 24, 2020

Report No.: TFF9H599 Quantity: 1PC Page Order/Pages: (P1/5) Ref. No.: NIL

Report Title: Curie Limited(U3104) Item: Mask

Address: No. 108, Ganhe Rd., Xitun Dist., Taichung City 407, Taiwan

Test Items	Test Results	Test Methods
Air Exchange Pressure (Pa/cm ²)	1	50.7
	2	49.1
	3	49.7
	4	51.3
	5	48.8
		EN 14683:2019 Annex C


Note: Sample description is given by the client: Brandname: Curie

Model: Curie KV99 Face Mask

Hong Kong address: Unit C, 23F, Tsun Tung Factory Bldg 38-40

Chai Wan Kok St, Tsuen Wan, NT Hong Kong

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.



- Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.
- 2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.
- 3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of
Taiwan Textile Research Institute

Jui-hung Kuo

Director,
Department of Testing and
Certification

Department of Testing and Certification, Taiwan Textile Research Institute
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)
Tel : +886-2-22670321 ext. 7107, 7110
Fax : +886-2-22675108, +886-2-22689839

Synthetic Blood Penetration Pressure in ASTM F2101 / ASTM F1862 / EN14683

Proven that KV99 mask can effectively intercept fluid (160mmHg) from penetration, complying ASTM F2101 Level 3 and EN14683 Type IIR



TEST REPORT TUCHENG

Date: Aug. 07, 2020 Date of Receipt: Jul. 27, 2020

Report No.: TAG9G706 Quantity: 1PC Page Order/Pages: (P2/5) Ref. No.: NIL

Report Title: Curie Limited(U3104) Item: Mask

Address: No. 108, Ganhe Rd., Xitun Dist., Taichung City 407, Taiwan

Test Items		Test Results	Test Methods
Synthetic Blood Penetration Pressure:160 mmHg	1	None Seen	ASTM F2100-2019 9.4
	2	None Seen	ASTM F1862-2017
	3	None Seen	
	4	None Seen	
	5	None Seen	
	6	None Seen	
	7	None Seen	
	8	None Seen	
	9	None Seen	
	10	None Seen	
	11	None Seen	
	12	None Seen	
	13	None Seen	
	14	None Seen	
	15	None Seen	
	16	None Seen	

Note: Sample description is given by the client: Brandname: Curie

Model: Curie KV99 Face Mask

Hong Kong address: Unit C, 23F, Tsun Tung Factory Bldg 38-40

Chai Wan Kok St, Tsuen Wan, NT Hong Kong

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.

Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.

2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.

3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of
Taiwan Textile Research Institute

Jui-hung kao

Director,
Department of Testing and
Certification

Department of Testing and Certification, Taiwan Textile Research Institute
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)
Tel : +886-2-22670321 ext. 7107, 7110
Fax: +886-2-22675108, +886-2-22689839



TEST REPORT TUCHENG

Date: Aug.31,2020 Date of Receipt: Aug.24,2020

Report No.: TFF9H599 Quantity: 1PC Page Order/Pages: (P2/5) Ref. No.: NIL

Report Title: Curie Limited(U3104) Item: Mask

Address: No. 108, Ganhe Rd., Xitun Dist., Taichung City 407, Taiwan

Test Items		Test Results	Test Methods
Synthetic Blood Penetration Pressure:120 mmHg (16.0 kPa)	1	None Seen	EN 14683:2019
	2	None Seen	ISO 22609:2004
	3	None Seen	
	4	None Seen	
	5	None Seen	
	6	None Seen	
	7	penetration	
	8	None Seen	
	9	None Seen	
	10	None Seen	
	11	None Seen	
	12	None Seen	
	13	None Seen	
	14	None Seen	
	15	None Seen	
	16	None Seen	

Note: Sample description is given by the client: Brandname: Curie

Model: Curie KV99 Face Mask

Hong Kong address: Unit C, 23F, Tsun Tung Factory Bldg 38-40

Chai Wan Kok St, Tsuen Wan. NT Hong Kong

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.

Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.

2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.

3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of
Taiwan Textile Research Institute

Jui-hung kao

Director,
Department of Testing and
Certification

Department of Testing and Certification, Taiwan Textile Research Institute
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)
Tel : +886-2-22670321 ext. 7107, 7110
Fax : +886-2-22675108 , +886-2-22689839



TEST REPORT TUCHENG

Date: Aug. 31, 2020 Date of Receipt: Aug. 24, 2020

Report No.: TFF9H599 Quantity: 1PC Page Order/Pages: (P3/5) Ref. No.: NIL

Report Title: Curie Limited(U3104) Item: Mask

Address: No. 108, Ganhe Rd., Xitun Dist., Taichung City 407, Taiwan

Test Items		Test Results	Test Methods
Synthetic Blood Penetration	17	None Seen	EN 14683:2019
Pressure:120 mmHg (16.0 kPa)	18	None Seen	ISO 22609:2004
	19	None Seen	
	20	None Seen	
	21	None Seen	
	22	None Seen	
	23	None Seen	
	24	None Seen	
	25	None Seen	
	26	None Seen	
	27	None Seen	
	28	None Seen	
	29	None Seen	
	30	None Seen	
	31	None Seen	
	32	None Seen	

Note: Sample description is given by the client: Brandname: Curie

Model: Curie KV99 Face Mask

Hong Kong address: Unit C, 23F, Tsun Tung Factory Bldg 38-40

Chai Wan Kok St, Tsuen Wan. NT Hong Kong

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.

Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.

2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.

3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of
Taiwan Textile Research Institute

Jui-hung kao

Director,
Department of Testing and
Certification

Department of Testing and Certification, Taiwan Textile Research Institute
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)
Tel : +886-2-22670321 ext. 7107, 7110
Fax : +886-2-22675108 , +886-2-22689839



TEST REPORT TUCHENG

Date: Aug.07,2020 Date of Receipt: Jul.27,2020

Report No.: TAG9G706 Quantity: 1PC Page Order/Pages: (P3/5) Ref. No.: NIL

Report Title: Curie Limited(U3104) Item: Mask

Address: No. 108, Ganhe Rd., Xitun Dist., Taichung City 407, Taiwan

Test Items		Test Results	Test Methods
Synthetic Blood Penetration Pressure:160 mmHg	17	None Seen	ASTM F2100-2019 9.4
	18	None Seen	ASTM F1862-2017
	19	None Seen	
	20	None Seen	
	21	None Seen	
	22	None Seen	
	23	None Seen	
	24	None Seen	
	25	None Seen	
	26	None Seen	
	27	None Seen	
	28	None Seen	
	29	None Seen	
	30	None Seen	
	31	None Seen	
	32	None Seen	

Note: Sample description is given by the client: Brandname: Curie

Model: Curie KV99 Face Mask

Hong Kong address: Unit C, 23F, Tsun Tung Factory Bldg 38-40

Chai Wan Kok St, Tsuen Wan. NT Hong Kong

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.

Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.

2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.

3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of
Taiwan Textile Research Institute

Jui-hung kao

Director,
Department of Testing and
Certification

Department of Testing and Certification, Taiwan Textile Research Institute
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)
Tel : +886-2-22670321 ext. 7107, 7110
Fax : +886-2-22675108, +886-2-22689839



TEST REPORT TUCHENG

Date: Aug. 31, 2020 Date of Receipt: Aug. 24, 2020

Report No.: TFF9H598 Quantity: 1PC Page Order/Pages: (P4/5) Ref. No.: NIL

Report Title: Curie Limited(U3104) Item: Mask

Address: No. 108, Ganhe Rd., Xitun Dist., Taichung City 407, Taiwan

Test Items		Test Results	Test Methods
Bacterial Filtration	1	99.9	ASTM F2100-2019 9.1 ASTM F2101-2019
Efficiency (BFE)(%)	2	99.9	
Staphylococcus aureus	3	> 99.9	
ATCC 6538	4	99.9	
	5	99.9	

Note: Control average: 2640 CFU.

Note: Mean particle size: 2.8 μ m.

Note: Testing side: outside of specimen.

Note: Testing area: 39.5 cm².

Note: Flow rate : 28.3 L/min.

Note: Sample description is given by the client: Brandname: Curie

Model: Curie KV99 Face Mask

Hong Kong address: Unit C, 23F, Tsun Tung Factory Bldg 38-40

Chai Wan Kok St, Tsuen Wan, NT Hong Kong

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.

Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.

2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.

3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of
Taiwan Textile Research Institute

Jui-hung kao

Director,
Department of Testing and
Certification

Department of Testing and Certification, Taiwan Textile Research Institute
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)
Tel : +886-2-22670321 ext. 7107, 7110
Fax : +886-2-22675108 , +886-2-22689839



TEST REPORT TUCHENG

Date: Aug.31,2020 Date of Receipt: Aug.24,2020

Report No.: TFF9H599 Quantity: 1PC Page Order/Pages: (P4/5) Ref. No.: NIL

Report Title: Curie Limited(U3104) Item: Mask

Address: No. 108, Ganhe Rd., Xitun Dist., Taichung City 407, Taiwan

Test Items		Test Results	Test Methods
Bacterial Filtration	1	> 99.9	EN 14683:2019 Annex B
Efficiency (BFE)(%)	2	99.9	
Staphylococcus aureus	3	99.9	
ATCC 6538	4	> 99.9	
	5	99.9	

Note: Control average: 2640 CFU.

Note: Mean particle size: 2.8 μ m.

Note: Testing side: outside of specimen.

Note: Testing area: 39.5 cm².

Note: Flow rate : 28.3 L/min.

Note: Sample description is given by the client: Brandname: Curie

Model: Curie KV99 Face Mask

Hong Kong address: Unit C, 23F, Tsun Tung Factory Bldg 38-40

Chai Wan Kok St, Tsuen Wan, NT Hong Kong

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.

Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.

2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.

3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of
Taiwan Textile Research Institute

Jui-hung kao

Director,
Department of Testing and
Certification

Department of Testing and Certification, Taiwan Textile Research Institute
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)
Tel : +886-2-22670321 ext. 7107, 7110
Fax : +886-2-22675108 , +886-2-22689839



TEST REPORT TUCHENG

Date: Aug. 07, 2020 Date of Receipt: Jul. 27, 2020

Report No.: TFF9G707 Quantity: 1PC Page Order/Pages: (P5/6) Ref. No.: NIL

Report Title: Curie Limited(U3104) Item: Mask

Address: No. 108, Ganhe Rd., Xitun Dist., Taichung City 407, Taiwan

Test Items		Test Results	Test Methods
Microbial cleanliness (cfu/g)	1	13.7	EN 14683:2019
	2	10.6	EN ISO 11737-1:2018
	3	4.4	
	4	10.1	
	5	11.0	

Note: Sample description is given by the client: Brandname: Curie

Model: Curie KV99 Face Mask

Hong Kong address: Unit C, 23F, Tsun Tung Factory Bldg 38-40

Chai Wan Kok St, Tsuen Wan. NT Hong Kong

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.

Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.

2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.

3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of
Taiwan Textile Research Institute

Jui-hung Kuo

Director,
Department of Testing and
Certification

Department of Testing and Certification, Taiwan Textile Research Institute
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)
Tel : +886-2-22670321 ext. 7107, 7110
Fax : +886-2-22675108 , +886-2-22689839

Bacterial Filtration Efficiency (BFE) in ASTM F2101

Proven that Curie technology can effectively filter bacteria (>99%)



政府創辦·多元創新
Government established · Diversified and innovative

TEST REPORT

Applicant: Curie Limited
Room C, 23/F,
Tsuen Tung Factory Building,
38-40 Chai Wan Kok Street,
Tsuen Wan,
New Territories,
Hong Kong

Report number: IRITS202005150001

Date: 15 May 2020

Attn.: Aldrin Or

Sample Description as Declared:

No. of Sample: TWO (2) pieces of received material in zipper bag packaging
Sample Description: Curie Ultrahigh- Efficiency Viral Filter
Colour: White
Date Received: 8 May 2020
Testing Period: 9 – 14 May 2020
Tests Conducted: As requested by the Applicant, with the details as follow:

Testing Summary: The sample being tested was conditioned for a minimum of 4 hour at 21 ± 5 °C and relative humidity of 65 ± 5 %. The bacterial filtration efficiency (BFE) test was performed by applying a spray of challenge bacterium *Staphylococcus aureus* in peptone water (approximately 2,200 colony forming units per spray) using a trigger sprayer. The sprayed aerosol was then drawn through the material being tested following by a tryptic soy agar plate under vacuum (flow rate: 100 Litres per minute). Number of *Staphylococcus aureus* colonies formed on the tryptic soy agar plate were counted after incubated at 37 ± 2 °C for 48 ± 4 hr. The BFE test procedure was modified from ASTM F2101: 2019.

For and on behalf of
Institute for Research in Innovative Technology & Sustainability
The Open University of Hong Kong

Dr. Eric Tung-po Sze
Director



香港公開大學
THE OPEN UNIVERSITY
OF HONG KONG

政府創辦・多元創新

Government established • Diversified and innovative

Report number: IRITS202005150001

Date: 15 May 2020


Results:

Test Sample Number

Test Sample Number	Bacterium Colonies Formed
#1	N.D. ^a
#2	N.D. ^a
Negative Control	N.D. ^a

^a None Detected (N.D.) – There were no detected bacterium colony of *Staphylococcus aureus* found.

Sample Photo:



<End of Test Report>

Standard Guide for Accelerated Ageing of Sterile Barrier Systems for Medical Devices in ASTM F1980-16
Bacterial Filtration Efficiency (BFE) in ASTM F2101

Proven that Curie technology can effectively filter bacteria (>99%), after conditioning KV99 masks in 120°C for 48 hours, to simulate storing in room temperature for 5 years



TEST REPORT

Applicant: Curie Limited
Room C, 23/F,
Tsuen Tung Factory Building,
38-40 Chai Wan Kok Street,
Tsuen Wan,
New Territories,
Hong Kong

Report number: IRITS2020007030001

Date: 3 July 2020

Attn.: Aldrin Or

Sample Description as Declared:

No. of Sample: TWO (2) pieces of composite material for face mask in zipper bag packaging
Curie KV99
Colour: White
Date Received: 15 June 2020
Testing Period: 16 – 24 June 2020
Tests Conducted: As requested by the Applicant, with the details as follow:

Testing Summary: The sample(s) were conditioned at an acceleration temperature of 120 °C for 48 hours, followed by pre-conditioning at a minimum of 4 hour at 21 ± 5 °C and relative humidity of 65 ± 5 %. Bacterial filtration efficiency (BFE) test was then performed by spraying the samples with an aerosol of challenge bacterium *Staphylococcus aureus* in peptone water using a nebulizer. The aerosol was then drawn through the samples following by a tryptic soy agar plate under vacuum (flow rate: 100 Litres per minute). Number of *Staphylococcus aureus* colonies formed on the tryptic soy agar plate were counted after incubated at 37 ± 2 °C for 48 ± 4 hr. The BFE test procedure was modified from ASTM F2101: 2019.

For and on behalf of
Institute for Research in Innovative Technology & Sustainability
The Open University of Hong Kong



Dr. Eric Tung-po Sze

Report number: IRITS2020007030001

Date: 3 July 2020

Results:

Test Sample Number	Bacterium Colonies Formed	Bacterial Filtration Efficiency
#1	N.D. ^a	> 99 %
#2	N.D. ^a	> 99 %
Negative Control	N.D. ^a	N/A ^b

^a None Detected (N.D.) – There were no detected bacterium colony of *Staphylococcus aureus* found

^b N/A – Not Applicable

Remark: The time and temperature selected for the acceleration conditioning were based on ASTM Standard F1980-16 Appendix X1. Accelerated aging of polymers, which are equivalent to five year of room-temperature (20 °C) aging, with an aging factor $Q_{10} = 2.0$.

Sample Photos:



<End of Test Report>



科技學院 *School of Science and Technology*

TEST REPORT

Applicant: Curie Limited
Room C, 23/F,
Tsuen Tung Factory Building,
38-40 Chai Wan Kok Street,
Tsuen Wan,
New Territories,
Hong Kong

Report number: IRITS2020007130001R1

Date: 23 July 2020

Attn.: Aldrin Or

Sample Description as Declared:

No. of Sample: ONE (1) piece of textile material in zipper bag packaging said to be RT-2007-T0430-DC020
Colour: White
Date Received: 21 May 2020
Testing Period: 2 – 10 July 2020
Tests Conducted: As requested by the Applicant to determine the antibacterial activity of the sample with reference to BS EN ISO 20743: 2013 Clause 8.2 Transfer method, with the following deviation:

- Shake-out the bacteria from specimens using peptone water instead of neutralizing solution.

For and on behalf of
Institute for Research in Innovative Technology & Sustainability
The Open University of Hong Kong



Dr. Eric Tung-po Sze
Director



Report number: IRITS2020007130001R1

Date: 23 July 2020

Results:

Specimen	Conditions	Number of bacteria ^a (CFU per specimen)
#1	Shake-out before incubation	0
#2	Shake-out after incubation	0

^a1 millilitre of an inoculum of *Staphylococcus aureus* with concentration of 1×10^6 CFU/ml to 3×10^6 CFU/ml was applied onto an agar plate in the transfer method, where each specimen was set on the agar surface and weigh down with a 200 g stainless-steel cylinder for $60 \text{ s} \pm 5 \text{ s}$ to transfer the microbial content. Incubation Measurement of the number of bacteria colonies was conducted in accordance with the plate count method specified in Annex C of BS EN ISO 20743:2013.

Opinion(s) and Interpretation(s): Based on the results obtained above, the specimens demonstrated effective antibacterial property to kill bacteria during transfer phase of the experiment.

Note: This Report replaces Report number IRITS2020007130001, which has been obsoleted.

<End of Test Report>



广微测
Gmicro Testing



中国认可
国际互认
检测
TESTING
CNAS L1747

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

REPORT FOR ANALYSIS

Report No.

2020FM20686R01E

Name of Sample

Curie Ultrahigh-Efficiency Viral Filter for KV-99

Applicant

Shenzhen Qianhai e-Cycle Trading Co.,Ltd.

Test Type

Entrustment Test

Address: Building 66, No.100 Central Xian Lie Road, Guangzhou, China

Postcode: 510070

Tel: +86 20 87137666

Fax : +86 20 87137668

Website : www.gddcm.com



广微测
Gmicro Testing



中国认可
国际互认
检测
TESTING
CNAS L1747

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

REPORT FOR ANALYSIS

Report No.: 2020FM20686R01E Verification Code: 03658924



Name of Sample	Curie Ultrahigh-Efficiency Viral Filter for KV-99	Test Type	Entrustment Test
Applicant	Shenzhen Qianhai e-Cycle Trading Co., Ltd.	Address	2/F Building B2, Yuntian Industrial Area, Xixiang Street, Baoan District, Shenzhen Guangdong, China
Sample Source	Submitted for Testing by the Applicant	Sample Quantity	260cm*2m
Spec and Lot No of Sample	40g; 1001	State and Characteristic	Flaky
Sample Received Date	2020-07-15	Test Completion Date	2020-07-28
Test Standard and Method	ISO 18184: 2014 (E)		
Item Tested	Antiviral activity test		
Test Conclusion	<p>The test data of the sample(s) is attached to the page(s) of this report.</p> <p>Issue Date: 2020-08-13 (Official Seal)</p>		
Remarks	<p>1. Manufacturer: Curie Limited. (provided by the applicant)</p> <p>2. Trademark: Curie; The date of production: 2020-06-01. (provided by the applicant)</p>		

Editor:

Chen Yingying

Verifier:

Li Sujuan

Approver:

Xie Xiaobao



广微测
Gmicro Testing



中国认可
国际互认
检测
TESTING
CNAS L1747

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

ANALYSIS AND TEST RESULT

Report No.: 2020FM20686R01E

Virus and host cell	No.	The logarithm of infectivity titre value immediate after inoculation of the reference specimen (lgTCID ₅₀ /bottle)	The logarithm of infectivity titre value after 2h contacting with the reference specimen (lgTCID ₅₀ / bottle)	The logarithm of infectivity titre value after 2h contacting with the test specimen (lgTCID ₅₀ / bottle)
H3N2 <i>Influenza A virus</i> Host cell: MDCK	1	7.05	6.50	2.10
	2	6.97	6.63	2.30
	3	7.10	6.59	2.30
lgTCID ₅₀ / bottle Average		7.04	6.57	2.33
Logarithm of antiviral activity		4.34		
Antiviral activity rate (%)		99.99		
(Blank below)				



Report №.: 2020FM20686R01E

Notice Items

1. The Test report is invalid if not affixed with Authorized Stamp of Test and Paging Seal.
2. The Test report is invalid without signature of verifier and approver.
3. The Test report is invalid if being supplemented, deleted or altered.
4. Without prior written permission, the report cannot be reproduced, except in full.
5. Unless otherwise stated, the results shown in this test report refer only to the sample(s) submitted.
6. Any dispute of the report must be raised to the testing body within 15 days after the report is received, exceeding which the dispute will not be accepted.
7. For the tested sample(s) submitted by the applicant, the sample information in the test report is provided by the applicant and the laboratory is not responsible for its authenticity.

Determination of alkylphenols (AP) of Textile Products EN ISO 21084:2019

Proven that Alkylphenols (AP) is not detected from Curie technology

Detection and Determination of Alkylphenol Ethoxylates (APEO) of Textile Products EN ISO 18254:2016

Proven that Alkylphenol Ethoxylates (APEO) is not detected from Curie technology



Technical Report: **(5220)210-0555**

August 5, 2020

Page 3 of 8

TEST RESULT

Alkylphenols (AP) Content Test

Test Method I : For Textile & Leather:
EN ISO 21084

Test Method II : For Polymers and other materials:
Organic solvent extraction, analysis with reference to EN ISO 21084.

Tested Item(s) : I001 White fabric with transparent adhesive

Maximum Limit:	100 mg/kg (sum)
-----------------------	------------------------

Tested Item(s)	Test Method	Result			Conclusion
		Detected Analyte(s)	Conc.	Unit	
I001	I	ND	ND	mg/kg	PASS

Note:

ND = Not detected

">" = More than

Conc. = Concentration

ppm = part(s) per million = mg/kg

mg/kg = milligram(s) per kilogram

Reporting Limit (mg/kg) : Sum (OP & NP) : 10

Remark:

- The list of alkylphenols is summarized in table of Appendix.

Alkylphenol Ethoxylates (APEO) Content Test

Test Method I : For Textile and Other materials:
With reference to EN ISO 18254-1, analysis by Liquid Chromatograph Mass Spectrometer (LC-MS)

Test Method II : For Leather:
With reference to EN ISO 18218-1 and EN ISO 18254-1

Tested Item(s) : I001 White fabric with transparent adhesive

Maximum Limit:	Others: 100 mg/kg (Sum) Recycled materials: 1000 mg/kg (Sum)
-----------------------	---

Tested Item(s)	Test Method	Result			Conclusion
		Detected Analyte(s)	Conc.	Unit	
I001	I	ND	ND	mg/kg	PASS

Note:

ND = Not detected

">" = More than

Conc. = Concentration

ppm = part(s) per million = mg/kg

mg/kg = milligram(s) per kilogram

Reporting Limit (mg/kg) : Sum (OPEOs & NPEOs) : 20

Remark:

- The list of alkylphenol ethoxylates is summarized in table of Appendix.

Determination of Formaldehyde - Free and Hydrolysed Formaldehyde of Textile Products

EN ISO 14184:2011 / JIS L 1041

Proven that Formaldehyde is not detected from Curie technology, and it reach safety level for Type 1 – Baby < 36 Months



Technical Report: **(5220)210-0555**

August 5, 2020

Page 4 of 8

TEST RESULT

Formaldehyde Content Test

- Test Method I** : Textiles & Other Materials: EN ISO 14184-1
Test Method II : Leather: ISO 17226:2 and/or ISO 17226-1
Test Method III : Glue & 0-3 years old products: JIS L 1041: 2011 Method A
Test Method IV : Carpets and mats: GB 18587 Grade B.
Tested Item(s) : I001 White fabric with transparent adhesive

Maximum Limit:	Type I	For baby < 36 months: 16 mg/kg			
	Type II	Others: 75 mg/kg			
	Type III	Carpets and mats : < 0.050 mg/m²/h			
Tested Item(s)	Type	Test Method	Result	Unit	Conclusion
I001	I	III	ND	mg/kg	PASS

Note:

ND = Not detected

“>” = More than

ppm = part(s) per million = mg/kg

mg/kg = milligram(s) per kilogram

Reporting Limit : 16 mg/kg (Method I, II, III), 0.050 mg/m²/h (Method IV)

Determination of Tetrachlorophenol-, Trichlorophenol-, Dichlorophenol-, Monochlorophenol-Isomers and Pentachlorophenol Content of Textile Products

DIN EN ISO 17070:2015 / 64 LFBG B 82.02-08 (Modified)

Proven that Chlorophenols Content and Ortho-Phenylphenol (OPP) are not detected from Curie technology



Technical Report: (5220)210-0555

August 5, 2020

Page 5 of 8

TEST RESULT

Chlorophenols Content Test

Test Method : With reference to ISO 17070:2015 or 64 LFBG B 82.02-08 (Modified). Potassium hydroxide extraction, derivatisation and analysis by Gas Chromatograph Mass Spectrometer (GC-MS).

Tested Item(s) : I001 White fabric with transparent adhesive

Maximum Limit:	0.5 mg/kg (Each)			
Tested Item(s)	Result			Conclusion
	Detected Analyte(s)	Conc.	Unit	
I001	ND	ND	mg/kg	PASS

Note:

ND = Not detected

“>” = More than

ppm = part(s) per million = mg/kg

mg/kg = milligram(s) per kilogram

Reporting Limit (mg/kg) : Each : 0.5

Remark:

- The list of chlorophenols is summarized in table of Appendix.

Ortho-phenylphenol (OPP) Test

Test Method : 1 M KOH extraction, 16 hours at 90 degrees C, derivatization and analysis § 64 LFBG B 82.02-08 or DIN EN ISO 17070:2015.

Tested Item(s) : I001 White fabric with transparent adhesive

Maximum Limit:	1000 mg/kg		
Tested Item(s)	Result	Unit	Conclusion
I001	ND	mg/kg	PASS

Note:

ND = Not detected

“>” = More than

ppm = part(s) per million = mg/kg

mg/kg = milligram(s) per kilogram

Reporting Limit (mg/kg) : 100

Determination of the Phthalate Content of Textile Products EN ISO 14389: 2014 / US CPSC-CH-C1001-09.4

Proven that Phthalate Content are not detected from Curie technology

Phthalates Content Test

Test Method I : Textile: CPSC-CH-C1001-09.4 and EN ISO 14389: 2014.

Test Method II : Others: CPSC-CH-C1001-09.4, analysis by Gas Chromatograph Mass Spectrometer (GC-MS).

Tested Item(s) : I001 White fabric with transparent adhesive

Maximum Limit:	500 mg/kg (Each) 1000 mg/kg (Sum)
-----------------------	--

Tested Item(s)	Test method	Result			Conclusion
		Detected Analyte(s)	Conc.	Unit	
I001	I	ND	ND	mg/kg	PASS

Note:

ND = Not detected

ppm = part(s) per million = mg/kg

Reporting Limit (mg/kg) : Each : 50

“>” = More than

mg/kg = milligram(s) per kilogram

Conc. = Concentration

Remark:

- The list of phthalates is summarized in table of Appendix.